Service Software Page 1 of 13

Quality Assurance (04-May-2020, 10:33:03, #141814)

Coil Check 16Ch_Breast_H_Sen Success (04-May-2020, 10:33:04 » 04-May-2020, 10:38:50)

Comments (0)

#	Timestamp	Comment	
No	No comments for this function.		

Summary

Summary

	Status
098_16Ch_Breast_BL_H_Sen_Tra	OK
098_16Ch_Breast_BR_H_Sen_Tra	OK
Program Result	Success

Details


Protocol: 098_16Ch_Breast_BL_H_Sen_Tra

Coil Info

Coil ID	Serial Number	Part Number
16Ch_Breast_H_Sen	60000032	10606804

Coil: BL1_BL2_BL3_BL4_ML1_ML2_ML3_ML4

Signal



Parabola Fit		
Value Min		
62.8	50.0	

Service Software Page 2 of 13

Brightness


Parabola Fit			
Value	Min	Max	
650	300	1400	

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BL1

Signal



S/N

Parabola Fit		
Value	ie Min	
74.0	50.0	

Brightness


Parabola	Fit
Ę	595

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BL2

Service Software Page 3 of 13



S/N

Parabola Fit		
Value	Min	
59.4	38.0	

Brightness


Parabola	Fit
5	545

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BL3

Signal



Parabola Fit		
Value	Min	
57.2	35.0	

Service Software Page 4 of 13

Brightness


Parabola Fit
582

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BL4

Signal



S/N

Parabola Fit		
Value	Min	
70.3	50.0	

Brightness


Parabola	Fit
5	574

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: ML1

Service Software Page 5 of 13



S/N

Parabola Fit		
Value	Min	
54.7	50.0	

Brightness


Parabola	Fit
3	395

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: ML2

Signal



Parabola Fit		
Value	Min	
48.5	45.0	

Brightness


Parabola Fit
419

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: ML3

Signal



S/N

Parabola Fit		
Value	Min	
44.8	35.0	

Brightness


Parabola	Fit
	381

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: ML4

Service Software Page 7 of 13



S/N

Parabola Fit		
Value	Min	
43.0	33.0	

Brightness

Parabola	Fit
4	109

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V


Protocol: 098_16Ch_Breast_BR_H_Sen_Tra

Coil Info

Coil ID	Serial Number	Part Number
16Ch_Breast_H_Sen	60000032	10606804

Coil: BR1_BR2_BR3_BR4_MR1_MR2_MR3_MR4

Service Software Page 8 of 13



S/N

Parabola Fit		
Value	Min	
62.5	50.0	

Brightness


Parabola Fit		
Value	Min	Max
656	300	1400

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BR1

Signal



3/19		
Parabola Fit		
Value	Min	

66.9 50.0

Brightness


Parabola Fit 511

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BR2

Signal



S/N

Parabola Fit	
Value	Min
51.2	38.0

Brightness


Parabola Fit 509

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	>

Coil: BR3

Service Software Page 10 of 13



S/N

Parabola Fit	
Value	Min
57.0	35.0

Brightness


Parabola	Fit
5	72

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: BR4

Signal



Parabola Fit		
Value	Min	
67.1	50.0	

Brightness


Parabola Fit	:
503	,

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: MR1

Signal



S/N

Parabola Fit	
Value	Min
66.3	50.0

Brightness


Parabola	Fit
4	184

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: MR2

Service Software Page 12 of 13



S/N

Parabola Fit		
Value	Min	
59.1	45.0	

Brightness


Parabola	Fit
5	500

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: MR3

Signal



Parabola Fit		
Value	Min	
53.2	35.0	

Service Software Page 13 of 13

Brightness


Parabola	Fit
4	187

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V

Coil: MR4

Signal



S/N

Parabola Fit	
Value	Min
49.2	33.0

Brightness

Parabola Fit 529

Homogeneity and Transmitter Amplitude

Variable	Value	Unit
Transmitter Amplitude	305	V