



THERMOMETER CALIBRATION LOG

SQF NO. 11.2.3

Thermometer ID:					THERMOMETER ACTUAL READING				
NO.	DATE	TIME	THERMOMETER ID	HOT	COLD	PASS/FAIL	CHECKED BY	CORRECTIVE ACTION	

Critical Limits: Each thermometer must be within +/- 2 OF of the certified thermometer.

Hot Calibration: Verify the boiling point of water based on the location's altitude. (Jarrell, TX BP = 210.4OF or 99OC)

Corrective Action: Difference of >/- 2 OF results in recalibration of thermometer.

Broken thermometers will be taken out of service and documented in the corrective action column. (When found during the calibration process)

Monitoring: Each thermometer will be calibrated on a daily basis.

Verification: Owner or designated employee will review this record prior to shipment of product.

Verified By:

Date:

Time: