

Electrometer charge calibration report for file: exp112847

```
CRC32 file validation = TRUE (ok.lines=43, not.ok.lines=0)
Consistency between index and number of lines in data file = TRUE
> Notes made during the analysis:
42-step procedure
> End of notes.
```



item	value
File	exp112847
Cal. note	Cal-session-15 XFBK SC1000SN806170
Model (electrometer)	'KEITHLEYINSTRUMENTSINCMODEL6517B-A12/700x'
SN (electrometer)	1341729
Range (electrometer)	2e-06
Range of charge supplied (min / max)	-2.0 nC / 2.0 nC (max current = 0.130 nA)
Cable (data file)	'Keithley-triax'
Capacitor id (data file)	'1 nF 1404A-SN-D7-16431384'
Capacitor note (script)	1404A-SN-D7-16431384
Capacitor value (script)	1.000028 nF
Voltmeter data: base value before step	V.ext.read4 V.ext.read5
Voltmeter data: main reading after step	V.ext.read6 V.ext.read7
Electrometer data: base value before step	nC.read4 nC.read5
Electrometer data: main reading after step	nC.read6 nC.read7
Leakage correction model	uniform.leak.model
Voltage burden	0.0 uV
Measurement time: start - stop	'12-07-2018 - 02:56:09' - '12-07-2018 - 04:25:09'
Duration	89.01 min
Lab. temperature	22.15 degC
Lab. humidity	47.0 %RH
Lab. pressure	1014.28 hPa
Ramp time / Wait.before.first.read	15 s / 15 s
Wait.leakage / Wait.between.readings / Total step duration	5 s / 1 s / 98.4 s
Initial charge reading (raw) after reset	-5.5800e-01 pC sd = 1.2400e-01 pC N = 42
End charge reading (raw) after full charge return	1.010 pC sd = 1.423 pC N = 42 Min = -3.1 pC Max = 5.5 pC
Net charge (leakage corrected etc.) for zero-V step	0.115 pC sd = 1.099 pC N = 4
Leakage current (absolute values)	Before: Median= 0.042 pA After: Median= 0.033 pA Max= 0.101 pA
k.elec.all (all data pooled; stat.limit = 25 %)	-1.000474 nC/nC sd = -0.038 % N = 16
k.elec.pos (positive charge only)	-1.000473 nC/nC sd = -0.040 % N = 8
k.elec.neg (negative charge only)	-1.000475 nC/nC sd = -0.039 % N = 8
Polarity ratio: k.pol = k.elec.pos / k.elec.neg	0.999998
Polarity difference : k.elec.pos - k.elec.neg (ANOVA)	0.000002 + - 0.000197 p = 0.991
Reference polarity for k.non.lin coeff	Negative (k.elec.neg)
Leakage correction uncert. = u.k.elec.model	-0.0006 %

=	DI	ſ
	*	1
	₹	₹

item		value
File	exp112847	
Cal. note	Cal-session-15 XFBK SC1000SN806170	
Model (electrometer)	'KEITHLEYINSTRUMENTSINCMODEL6517B-A12/700x'	
SN (electrometer)	1341729	
Range (electrometer)	2e-06	
Range of charge supplied (min / max)	-2.0 nC / 2.0 nC (max current = 0.130 nA)	
UPARM.charge.u.base.abs	0.000000	
UPARM.charge.u.base.pct	0.043000	
UPARM.charge.u.minimum.step.abs	0.000000	
UPARM.charge.u.minimum.step.pct	0.050000	
UPARM.charge.u.voltage.uV	40.000000	
UPARM.charge.u.transfer.time.s	5.000000	
UPARM.AGGREGATE.charge.u.base.abs	0.000000	
UPARM.AGGREGATE.charge.u.base.pct	0.043000	
UPARM.AGGREGATE.charge.u.minimum.abs	0.000000	
UPARM.AGGREGATE.charge.u.minimum.pct	0.050000	
UPARM.charge.u.base.abs	0.000000	
k.elec.all (all data pooled; stat.limit = 25 %)	-1.000474 nC/nC sd = -0.038 % N = 16	
k.elec.pos (positive charge only)	-1.000473 nC/nC sd = -0.040 % N = 8	
k.elec.neg (negative charge only)	-1.000475 nC/nC sd = -0.039 % N = 8	
UAGGREGATE.k.elec.all	0.003440 (k=2)	
UAGGREGATE.k.elec.pos	0.003447 (k=2)	
UAGGREGATE.k.elec.neg	0.003260 (k=2)	
Leakage correction uncert. = u.k.elec.model	-0.0006 %	

Electrometer charge calibration (3 digits): 'KEITHLEYINSTRUMENTSINC.-MODEL6517B-A12/700x'SN1341729 (exp112847)

Non-linearity correction factors (reference = k.elec.neg; Negative):

V.select	V.ref	pA.Before	pA.After	nC.adj	nC.net	nC.ref	k.elec.step	N	u.k.elec.step	k.non.lin
-2	-2.001	-0.02	0.02	0.001	2.000	-2.001	-1.000	1	0.001	1
-1.8	-1.801	-0.04	0.01	0.001	1.800	-1.801	-1.000	2	0.001	0.9999
-1.6	-1.601	0.03	0.00	-0.000	1.600	-1.601	-1.001	2	0.001	1.0001
-1.4	-1.401	0.03	0.02	0.001	1.400	-1.401	-1.000	2	0.001	0.9997
-1.2	-1.201	-0.03	0.01	0.000	1.199	-1.201	-1.001	2	0.001	1.0007
-1	-1.000	0.01	0.07	0.001	1.000	-1.000	-1.001	2	0.001	1.0002
-0.8	-0.800	-0.10	0.06	0.001	0.800	-0.800	-1.000	2	0.001	0.9999
-0.6	-0.600	0.04	-0.01	0.001	0.600	-0.600	-1.000	2	0.002	0.9994
-0.4	-0.400	0.01	0.04	0.001	0.400	-0.400	-1.002	2	0.002	1.0013
-0.2	-0.200	-0.01	-0.03	0.000	0.200	-0.200	-1.001	2	0.004	1.0002
0.2	0.200	0.01	0.01	0.001	-0.200	0.200	-1.000	2	0.004	0.9999
0.4	0.400	0.08	-0.03	0.001	-0.401	0.400	-0.999	2	0.002	0.9984
0.6	0.600	-0.06	0.07	0.001	-0.600	0.600	-1.001	2	0.002	1.0007
0.8	0.800	-0.02	0.03	0.001	-0.800	0.800	-1.001	2	0.001	1.0003
1	1.000	0.06	0.08	0.001	-1.000	1.000	-1.000	2	0.001	0.9997
1.2	1.200	0.01	0.03	0.000	-1.200	1.200	-1.000	2	0.001	0.9996
1.4	1.401	0.01	0.01	0.001	-1.400	1.401	-1.000	2	0.001	0.9997
1.6	1.601	0.07	-0.00	0.001	-1.600	1.601	-1.001	2	0.001	1.0001
1.8	1.801	0.04	0.01	0.001	-1.800	1.801	-1.001	2	0.001	1.0001
2	2.001	-0.08	0.05	-0.000	-2.000	2.001	-1.000	1	0.001	0.9997

DTU


Electrometer charge calibration (4 digits): 'KEITHLEYINSTRUMENTSINC.-MODEL6517B-A12/700x'SN1341729 (exp112847)
Non-linearity correction factors (reference = k.elec.neg; Negative):


V.select	V.ref	pA.Before	pA.After	nC.adj	nC.net	nC.ref	k.elec.step	N	u.k.elec.step	k.non.lin
-2	-2.0008	-0.024	0.020	0.0010	1.9999	-2.0008	-1.0005	1	0.0008	1
-1.8	-1.8007	-0.040	0.008	0.0012	1.8000	-1.8008	-1.0004	2	0.0009	0.9999
-1.6	-1.6007	0.028	0.003	-0.0001	1.5997	-1.6007	-1.0006	2	0.0008	1.0001
-1.4	-1.4006	0.032	0.016	0.0005	1.4004	-1.4006	-1.0002	2	0.0009	0.9997
-1.2	-1.2005	-0.032	0.012	0.0003	1.1991	-1.2006	-1.0012	2	0.0009	1.0007
-1	-1.0005	0.009	0.074	0.0007	0.9998	-1.0005	-1.0007	2	0.0011	1.0002
-0.8	-0.8004	-0.095	0.055	0.0009	0.8001	-0.8004	-1.0004	2	0.0014	0.9999
-0.6	-0.6003	0.040	-0.015	0.0009	0.6004	-0.6003	-0.9998	2	0.0016	0.9994
-0.4	-0.4002	0.007	0.041	0.0011	0.3995	-0.4002	-1.0017	2	0.0024	1.0013
-0.2	-0.2001	-0.006	-0.028	0.0004	0.2000	-0.2001	-1.0007	2	0.0038	1.0002
0.2	0.2001	0.007	0.014	0.0010	-0.2000	0.2001	-1.0004	2	0.0044	0.9999
0.4	0.4001	0.080	-0.034	0.0008	-0.4006	0.4001	-0.9989	2	0.0024	0.9984
0.6	0.6002	-0.063	0.069	0.0007	-0.5995	0.6002	-1.0012	2	0.0017	1.0007
0.8	0.8003	-0.015	0.033	0.0005	-0.7997	0.8003	-1.0008	2	0.0012	1.0003
1	1.0004	0.057	0.075	0.0014	-1.0002	1.0004	-1.0002	2	0.0013	0.9997
1.2	1.2005	0.009	0.031	0.0001	-1.2004	1.2005	-1.0001	2	0.0009	0.9996
1.4	1.4005	0.014	0.010	0.0007	-1.4004	1.4006	-1.0001	2	0.0009	0.9997
1.6	1.6006	0.067	-0.004	0.0006	-1.5997	1.6006	-1.0006	2	0.0009	1.0001
1.8	1.8006	0.042	0.006	0.0009	-1.7997	1.8007	-1.0006	2	0.0009	1.0001
2	2.0007	-0.078	0.047	-0.0001	-2.0003	2.0007	-1.0002	1	0.0008	0.9997


Electrometer charge calibration: 'KEITHLEYINSTRUMENTSINC.-MODEL6517B-A12/700x'SN1341729 (exp112847) Details for individual steps (electrometer calibration):


step.i	V.select	V.ref	pA.Before	pA.After	pA.Return	nC.adj	nC.net	nC.ref	k.elec.step
1	0.0000	-0.0000	-0.096	0.013	0.005	0.0004	-0.0007	-0.0000	0.0155
2	-0.2000	-0.2001	0.069	-0.015	-0.040	0.0002	0.2000	-0.2001	-1.0003
3	-0.4000	-0.4002	0.040	0.026	-0.117	0.0005	0.4008	-0.4002	-0.9983
4	-0.6000	-0.6003	0.046	-0.032	0.048	0.0008	0.6003	-0.6003	-1.0000
5	-0.8000	-0.8004	-0.089	0.072	0.066	0.0012	0.8003	-0.8004	-1.0002
6	-1.0000	-1.0005	-0.014	0.101	-0.083	0.0012	1.0004	-1.0005	-1.0001
7	-1.2000	-1.2005	-0.061	0.043	0.109	0.0006	1.1990	-1.2006	-1.0013
8	-1.4000	-1.4006	0.023	-0.011	0.050	0.0010	1.3999	-1.4006	-1.0005
9	-1.6000	-1.6007	-0.022	-0.016	0.034	-0.0012	1.5998	-1.6007	-1.0006
10	-1.8000	-1.8007	0.046	-0.048	0.052	0.0014	1.7990	-1.8008	-1.0010
11	-2.0000	-2.0008	-0.024	0.020	0.061	0.0010	1.9999	-2.0008	-1.0005
12	-1.8000	-1.8007	-0.125	0.064	0.020	0.0010	1.8011	-1.8008	-0.9998
13	-1.6000	-1.6007	0.079	0.022	0.046	0.0010	1.5997	-1.6007	-1.0006
14	-1.4000	-1.4006	0.041	0.043	0.024	0.0001	1.4008	-1.4006	-0.9999
15	-1.2000	-1.2005	-0.003	-0.020	0.060	-0.0001	1.1992	-1.2006	-1.0011
16	-1.0000	-1.0005	0.032	0.048	0.020	0.0003	0.9992	-1.0005	-1.0013
17	-0.8000	-0.8004	-0.102	0.038	0.003	0.0006	0.7999	-0.8004	-1.0007
18	-0.6000	-0.6003	0.033	0.003	0.045	0.0010	0.6005	-0.6003	-0.9996
19	-0.4000	-0.4002	-0.026	0.056	0.096	0.0018	0.3982	-0.4002	-1.0051
20	-0.2000	-0.2001	-0.081	-0.041	0.029	0.0007	0.1999	-0.2001	-1.0011
21	0.0000	0.0000	0.028	-0.011	0.021	0.0002	0.0006	0.0000	0.0177
22	0.0000	-0.0000	-0.042	0.016	0.060	0.0004	-0.0009	-0.0000	0.0002
23	0.2000	0.2001	0.006	0.007	0.042	0.0011	-0.1991	0.2001	-1.0050
24	0.4000	0.4001	0.177	-0.082	-0.011	0.0004	-0.4003	0.4001	-0.9996
25	0.6000	0.6002	-0.037	0.059	0.063	0.0002	-0.6007	0.6003	-0.9992
26	0.8000	0.8003	0.017	0.033	-0.004	0.0009	-0.7994	0.8003	-1.0011
27	1.0000	1.0004	0.077	0.055	-0.097	0.0002	-1.0001	1.0004	-1.0003
28	1.2000	1.2005	0.043	0.034	0.040	0.0003	-1.2010	1.2005	-0.9995
29	1.4000	1.4005	0.005	-0.020	0.049	0.0006	-1.3994	1.4006	-1.0009
30	1.6000	1.6006	0.066	-0.008	-0.041	0.0008	-1.5993	1.6006	-1.0008
31	1.8000	1.8006	-0.005	0.035	-0.018	0.0017	-1.7994	1.8007	-1.0007
32	2.0000	2.0007	-0.078	0.047	-0.010	-0.0001	-2.0003	2.0007	-1.0002
33	1.8000	1.8006	0.089	-0.022	-0.101	0.0001	-1.7999	1.8007	-1.0005
34	1.6000	1.6006	0.067	-0.001	-0.042	0.0004	-1.6000	1.6006	-1.0004
35	1.4000	1.4005	0.023	0.040	0.003	0.0009	-1.4013	1.4006	-0.9994
36	1.2000	1.2005	-0.026	0.028	-0.103	-0.0000	-1.1998	1.2005	-1.0006
37	1.0000	1.0004	0.036	0.096	0.037	0.0026	-1.0003	1.0004	-1.0001
38	0.8000	0.8003	-0.048	0.034	0.040	0.0001	-0.8000	0.8003	-1.0005
39	0.6000	0.6002	-0.088	0.078	-0.017	0.0013	-0.5983	0.6002	-1.0032
40	0.4000	0.4001	-0.016	0.015	0.059	0.0012	-0.4008	0.4002	-0.9983
41	0.2000	0.2001	0.008	0.021	0.045	0.0010	-0.2009	0.2001	-0.9958
42	0.0000	0.0000	0.046	0.033	-0.004	0.0007	0.0014	0.0000	0.0013


Page 6 User: Claus E. Andersen Analyzed: 2020-03-21 17:05:22































k.non.lin vs. charge (exp112847) with k.elec.stat.limit = 25 %



Page 17 User: Claus E. Andersen Analyzed: 2020-03-21 17:05:22



This is the final page of the Electrometer charge calibration report for file: exp112847