

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/325416608>

# Suitability of the Low-Cost SDS011 Particle Sensor for Urban PM-Monitoring

Conference Paper · May 2018

CITATIONS

0

READS

6,584

9 authors, including:



**Matthias Budde**

Karlsruhe Institute of Technology

56 PUBLICATIONS 380 CITATIONS

[SEE PROFILE](#)



**Bernd Laquai**

Independent Citizen Scientist Stuttgart

41 PUBLICATIONS 107 CITATIONS

[SEE PROFILE](#)



**Norbert Streibl**

4 PUBLICATIONS 2 CITATIONS

[SEE PROFILE](#)



**Almuth Schwarz**

Karlsruhe Institute of Technology

6 PUBLICATIONS 3 CITATIONS

[SEE PROFILE](#)

Some of the authors of this publication are also working on these related projects:



Strategien zur Emissionsminderung von druckgepulsten Oberflächenfiltern [View project](#)



Innovative Ansätze zur Optimierung des Betriebsverhaltens von Oberflächenfiltern zur Gasreinigung [View project](#)

## SUITABILITY OF THE LOW-COST SDS011 PARTICLE SENSOR FOR URBAN PM MONITORING

Matthias Budde\* (1), Thomas Müller (2), Bernd Laquai (3), Norbert Streibl (3), Almuth D. Schwarz (4), Gregor Schindler (1), Till Riedel (1), Michael Beigl (1), Achim Dittler (4)

(1) Karlsruhe Institute of Technology (KIT), Pervasive Computing Systems / TECO, (2) Leibnitz Institute for Tropospheric Research (TROPOS), World Calibration Centre for Aerosol Physics (WCCAP), (3) Independent Researcher, (4) Karlsruhe Institute of Technology (KIT), Gas Particle Systems Group

In Particulate Matter (PM) monitoring, a paradigm shift towards incorporating distributed sensing approaches using low-cost sensors has begun [1]. In past research, early generations of low-cost particle sensors based on IR light scattering have been compared with official measurement stations, showing that these sensors can in principle capture the dynamics of ambient PM levels [2,3], but may suffer from low calibration stability [2], are unable to differentiate size classes [3], and may be susceptible to other sources of error [4]. Current low-cost sensor generations that rely on laser scattering claim to exhibit a better level of stability and feature internal digital processing in order to achieve more accurate results. While they are mostly designated as PM<sub>2.5</sub> sensors, some also output values for PM<sub>10</sub> and/or PM<sub>1</sub>.

As a representative of this class of sensors, we examine the SDS011 laser-scattering PM sensor [5]. It is already widely used in deployments around the world, e.g. in the German grassroots citizen science project “luftdaten.info” (<http://www.luftdaten.info>), in which volunteers have deployed hundreds of these sensors in urban areas. In previous work, co-location measurements between the SDS011 have already been performed [6], the results of which indicate that the sensor delivers adequate correlation under typical conditions (relative humidity of 20-50% and PM<sub>10</sub> mass concentrations < 20 µg/m<sup>3</sup>) but performs less well under other ambient conditions, especially high humidity. To further explore the sensor's data quality in-depth, we present the key influencing factors on measurement uncertainty of the low-cost sensor, along with a series of experiments to appropriately assess its potential and limitations:

- Investigation of the humidity influence and possibilities for its compensation.
- Comparison of the SDS011 sensor and a Welas2100 monitor using monodisperse aerosol of different sizes.
- Characterization of the mass distributions the SDS011 can capture, based on experiments with different generated particle spectra and using the Grimm 1.108 aerosol spectrometer as reference.
- Longer-term comparison (days) of 13 SDS011 and a Scanning Mobility Particle Sizer (SMPS) exposed to (1) ambient air, (2) artificial aerosol (ammonium sulfate) levels, and (3) black carbon/soot.

From the results of these experiments, we present the causes of the sensor's measurement uncertainty in our talk. We show that the sensor generally does not capture PM<sub>10</sub> satisfactorily and discuss under which conditions PM<sub>2.5</sub> readings reflect the ambient air quality adequately.

[Acknowledgements: Partially funded as part of BMBF project "Software Campus" (grant no. 01IS12051) and BMVI project "SmartAQnet" (grant no. 19F2003B).]

- [1] Snyder E. G., Watkins T. H., Solomon P. A., Thoma E. D., Williams R. W., Hagler G. S. W., Shelow D., Hindin D. A., Kilaru V. J., and Preuss P. W. (2013) The changing paradigm of air pollution monitoring. *Environmental Science & Technology*, vol. 47, no. 20.
- [2] Budde M., El Masri R., Riedel T., Beigl M. (2013) Enabling Low-Cost Particulate Matter Measurement for Participatory Sensing Scenarios, 12th International Conference on Mobile and Ubiquitous Multimedia (MUM 2013)
- [3] Holstius D. M., Pillariseti A., Smith K. R., Seto E. (2014) Field calibrations of a low-cost aerosol sensor at a regulatory monitoring site in California. *Atmos. Meas. Tech.*, 7, pp. 1121– 1131
- [4] Budde M., Köpke M., Beigl M. (2015). Robust, In-situ Data Reconstruction from Poisson Noise for Low-cost, Mobile, Non-Expert Environmental Sensing. *International Symposium on Wearable Computing (ISWC'15)*, pp. 179-182.
- [5] Nova Fitness Co., Ltd. (2015). SDS011 sensor. Version: V1.3, [Online]. <https://nettigo.pl/attachments/398> (visited on 01/24/2018).
- [6] LUBW (2017). Messungen mit dem Feinstaubsensor SDS011 - Ein Vergleich mit einem eignungsgeprüften Feinstaubanalysator. [Online]. <https://www4.lubw.baden-wuerttemberg.de/servlet/is/268831/> (visited on 01/24/2018).