



I:E Ratio

2:1 - 1:2

Respiratory Rate 5-40 bpm

Inspiratory pressure 40 cmH20

Expiratory pressure 0 - 20 cmH20

FiO2 20% -100%

Active Buzzer Alarm

Minute Ventilation (low/High) alarm Peak pressure, Low expiratory pressure Disconnection alarm Low expiratory pressure Oxygen concentration Non-rebreathing valves