

CHAI Applied Model Card

Name: BriefCase-Triage for Brain Aneurysm Developer: Aidoc	Inquires or to report an issue: www.aidoc.com/contact-support/ or email: info@aidoc.com
Release Stage: FDA Cleared, Commercial (K213721) Release Date: March 2022 Version: BA-01 Global Availability: US, EU (CE Marked), UK (UKCA Marked), Canada (MDEL), Australia (ARTG), New Zealand, Israel (AMAR), UAE and South Africa. Regulatory Approval: Please find attached the BA-01 510(k) summary .	
Summary: BriefCase is a radiological computer aided triage and notification software indicated for use in the analysis of head CT Angio (CTA) images. The device is intended to assist hospital networks and appropriately trained medical specialists in workflow triage by flagging and communication of suspected positive cases of Brain Aneurysm (BA) findings above 5 mm size. BriefCase uses an artificial intelligence algorithm to analyze images and flag suspect cases on a standalone desktop application in parallel to the ongoing standard of care image interpretation. The user is presented with notifications for suspect cases. Notifications include compressed preview images that are meant for informational purposes only and not intended for diagnostic use beyond notification. The device does not alter the original medical image and is not intended to be used as a diagnostic device. The results of BriefCase are intended to be used in conjunction with other patient information and based on professional judgment, to assist with triage/prioritization of medical images. Notified	Uses and Directions: <ul style="list-style-type: none">• Intended use and workflow: BriefCase is a radiological computer aided triage and notification software indicated for use in the analysis of head CT Angio (CTA) images. The device is intended to assist hospital networks and appropriately trained medical specialists in workflow triage by flagging and communication of suspected positive cases of Brain Aneurysm (BA) findings above 5 mm size.• Primary intended users: The Brain Aneurysm solution is intended to be used by appropriately trained medical specialists. Appropriately trained medical specialists for BA are Emergency physicians, radiologists, neurologists, neuro-surgeons, neuro-interventionalists, interventional radiologists, neuro intensive care physicians.• How to use: The user is presented with notifications for suspect cases. Notifications include compressed preview images that are meant for informational purposes only and not intended for diagnostic use beyond notification. The device does not alter the original medical image and is not intended to be used as a diagnostic device.• Targeted patient population: patients Adults or transitional adolescents aged 18 and older.• Cautioned out-of-scope settings and use cases: Notifications include compressed preview images that are meant for

clinicians are responsible for viewing full images per the standard of care.

Keywords: None

informational purposes only and not intended for diagnostic use beyond notification. The device does not alter the original medical image and is not intended to be used as a diagnostic device. Only for use in workflow triage, and only for use by appropriately trained medical specialists. Please refer to BriefCase-Triage User Guide ([UG0009](#)) for further information on warnings and precautions.

Warnings

- **Known risks and limitations:** Notified clinicians are responsible for viewing full images per the standard of care. Please refer to BriefCase-Triage User Guide ([UG0009](#)), label, and additional tutorial materials provided by the company for further information on risks and limitations.
- **Known biases or ethical considerations:** None
- **Clinical risk level:** Low, Class II (FDA) / IIa (EU)

Trust Ingredients

AI System Facts:

- **Outcome(s) and output(s):** Binary notification of suspected Brain Aneurysm cases. The desktop application feed displays incoming suspect cases. Hovering over the feed pops up a compressed, unannotated image that is captioned "not for diagnostic use." The radiologist prioritizes cases based on this additional info and reads the case as per standard of care.
- **Model type:** Custom-built 3D deep convolutional neural network.
- **Foundation models used in application:** N/A
- **Input data source:** The BriefCase-Triage imaging AI solutions are trained, tuned, and validated using diverse data sets. Specifically, Aidoc works with DICOM images, which capture anatomical and physiological data from medical imaging technologies like CT scans and X-rays. These data sets are sourced from a wide range of institutions, including community hospitals, academic centers, teleradiology providers, private clinics, and imaging centers across various global regions (e.g., east and west US, central US, EU, and other parts of the world).
- **Output/Input data type:** Input: Real World Digital imaging and communications in medicine (DICOM) images.

Inclusion criteria: computed tomography angiography (CTA) of the head. Scans performed on adults/transitional adults ≥ 18 years of age. Scans performed on CT scanners with 64 or greater number of detectors. Slice thickness; 0.5 – 2.5 mm axial slices.

Exclusion Criteria: All scans that are technically inadequate, including motion artifacts, severe metal artifacts, or an inadequate field of view.

Output: Binary notification of suspected cases of Brain Aneurysm

- **Development data characterization:** The Brain Aneurysm model has been trained, tuned, and validated on over 4,000 scans. The training and validation populations consist of a wide range of institutions and patient settings, including academic centers, community hospitals, teleradiology providers, private imaging centers, and clinics worldwide (specifically East, West, and Center of US, EU, Middle East, Oceania, Latin America and far east).

Analyzed data included key demographic elements such as age, gender, comorbidities, geographical location, patient settings, reason for exam, and technological characteristics (e.g., equipment manufacturer, slice thickness, and modality model distribution).

The time frame for the data used for the initial and retrained model construction and validation is 2017-2024.

- **Bias mitigation approaches:** Aidoc's AI solutions undergo rigorous training, validation, and monitoring to ensure robust and equitable performance across diverse patient populations. Bias mitigation and advancing care equity are highly prioritized throughout the design and development of Aidoc solutions, while working within the known constraints of limited accessibility to certain data elements.

Recognizing these constraints, Aidoc creates proxies to effectively mitigate risks associated with the absence of certain demographic details, when applicable.

Specifically, race, ethnicity, language, sexual orientation, gender identity, social determinants of health data, and health status assessment data are not accessible as part of the DICOM image.

To address potential biases, Aidoc employs diverse data sets from a wide range of institutions and patient settings, including academic centers, community hospitals, teleradiology providers, private imaging centers, and clinics worldwide (specifically East, West, and Center of US, EU, Middle East, Oceania, Latin America and far east).

Aidoc validates and measures the AI model across subpopulations to ensure a normal and predictable performance on them.

Additionally, key demographic elements are analyzed such as age, gender, comorbidities, geographical location, patient settings, reason for exam, and technological characteristics (e.g., equipment manufacturer, slice thickness, and modality model distribution);

Furthermore, Aidoc conducts multiple retraining throughout the lifecycle of each AI module to ensure increased generalizability across all the aforementioned parameters. While the exact size of Aidoc's training data sets is proprietary, over the model lifecycle, they may include up to millions of DICOM studies.

- **Ongoing Maintenance:** Aidoc's quality management system is ISO 13485 MDSAP and EU MDR certified, and 21 CFR Part 820 compliant, requiring structured and validated installation, servicing, and Post-Market Surveillance (PMS) activities.

Specifically, Aidoc's PMS plan outlines the systematic collection and analysis of data regarding the device's safety and performance throughout its lifecycle.

Data analysis methods, both qualitative and quantitative, are applied to the collected data. The results are documented in a PMS report and in a Periodic Safety Update Report (PSUR). In addition, Aidoc implements site-specific controls, ensuring high standards are consistently maintained across all operational locations. The current Brain Aneurysm model represents the first FDA clearance for Brain Aneurysm at Aidoc. Aidoc periodically retrains the algorithm on additional data and subjects the retrained versions to the required regulatory review and controls.

An expanded dataset enhances the model's generalization capabilities and potentially elevates the product's performance (improved or non-inferior to the previously cleared device's Time-to-

Notification and/or AUC). The inclusion of data from a broader and more diverse array of sources enhances the model's ability to perform effectively across a wider range of clinical and technical settings, particularly if any real-world data changes or if any data was underrepresented in the original dataset.

Continuous monitoring: As detailed in the previous answer, Aidoc has robust monitoring systems in place to track the AI performance and identify any emerging issues, such as AI drift. In addition, regular audits and assessments ensure the model continues to deliver fair and accurate prioritizations.

Furthermore, Aidoc remains committed to ongoing research and development to refine the AI and address any safety concerns that may arise. The Aidoc AI Monitoring team monitors algorithm performance 24/7 for the purpose of mitigating AI drift. The team monitors and considers aspects such as alert correctness, the timeliness of data available, data completeness, slice thickness, series correctness and relevancy, contrast phase, algorithm specificity, and algorithmic positive ratio.

- **Security and compliance environment practices or accreditations:** Aidoc's products are structured based on international quality, privacy, and security standards and frameworks, including ISO 13485, ISO 27001, ISO 27017, ISO 27018, ISO 27799, SOC 2 Type 2, Cyber Essentials, and C5. Aidoc follows a systematic approach, integrating these frameworks into Aidoc's Information Security Management System (ISMS) to ensure a holistic and proactive risk management strategy.

Aidoc's risk management program involves regular risk assessments, leveraging common methodologies, to identify, evaluate, and prioritize risks. Continuous monitoring, periodic audits, and assessments contribute to the dynamic nature of Aidoc's risk management program, allowing us to adapt to evolving threats and vulnerabilities. Aidoc regularly monitors its server's logs for unusual activity and implements intrusion detection and prevention systems, paying attention to unusual or suspicious activities such as unexpected errors, system slowdowns, or strange user access patterns. Furthermore, Aidoc conducts multiple retraining throughout the lifecycle of each AI module.

Additionally, Aidoc's Governance, Risk, and Compliance (GRC) Program is responsible for enterprise oversight and direction for all security governance activities; governance for risk-related activities including Risk Identification, Management, Mitigation, and Remediation, and Risk Assessments; and responsibility for the planning, execution, and adherence with Aidoc Security Policies and Procedures, legal, regulatory, and contractual requirements.

- **Transparency, Intelligibility, and Accountability mechanisms:** The company ensures Transparency, Intelligibility, and Accountability in Aidoc's medical devices via its product design and its Quality Management System (QMS). Some examples include (non-exhaustive list):
 - Explainability and transparency features by design;
 - Reporting and feedback mechanism directly via the device's UI and additional communication in different platforms;
 - User guide, labeling, and additional tutorial materials are available electronically at all times via the user community;
 - Robust user training.

Transparency Information:

- **Funding source of the technical implementation:** Aidoc's model creation is self-funded.
- **3rd Party Information:** N/A
- **Stakeholders consulted during design of intervention (e.g. patients, providers):** Physicians, health system administrators, and patient groups were engaged in the development of the

Key Metrics

Usefulness, Usability, and Efficacy		Fairness and Equity		Safety and Reliability	
Goal of metric: Evaluate the software's performance in identifying head CTs (with and without contrast) containing Brain Aneurysm from five study sites. 268 cases were selected, 96 positive cases and 172 negative cases (reports on images with Brain Aneurysm versus without Brain Aneurysm findings) included in the analysis.		Goal of metric: N/A		Goal of metric: The time-to-notification metric observed for the BriefCase software in the five medical centers was compared to the equivalent metric of prior predicate devices.	
Result: Sensitivity and specificity exceeded the 80% performance goal. Specifically, sensitivity was 88.5% (95% CI: 80.4%, 94.1%) and specificity was 89.5% (95% CI: 84.0%, 93.7%)	Interpretation: Primary endpoints were sensitivity and specificity with an 80% performance goal. Secondary endpoints were BriefCase time-to-notification compared to the predicate device, Positive Predictive Value (PPV), Negative Predictive Value (NPV), Positive Likelihood Ratio (PLR), and Negative Likelihood Ratio (NLR).	Result: Race, age, and gender distribution within the study data patient population was unavailable.	Interpretation: Race, age, and gender distribution within the study data patient population was unavailable.	Result: The standard of care time-to-exam-open (89.4 minutes; 95% CI: 56.0-122.7; Median 66.0, IQR 50.7) was significantly longer than the time-to-notification metric of the BriefCase device (4.2 minutes, 95% CI: 3.9-4.5; Median 4.2, IQR 1.8). The mean difference of 85.2 minutes (95% CI: 51.8-118.6; Median 63.1, IQR 50.1)	Interpretation: The time-to-notification results obtained for the subject BriefCase device showed improvement with regard to time savings to the standard of care review.
Test Type: Retrospective, blinded, multicenter study.		Test Type: Retrospective, blinded, multicenter study		Test Type: Retrospective, blinded, multicenter study	
Testing Data Description: Race, age, and gender distribution within the study data patient population was unavailable.		Testing Data Description: Race, age, and gender distribution within the study data patient population was unavailable.		Testing Data Description: The BriefCase time-to-notification includes the time to get the DICOM exam, de-identify it, upload it to the cloud, analyze and send a notification on a positive suspect case back to the desktop application. The BriefCase time-to-notification was measured for all True Positive cases (i.e., identified as positive both by the reviewers as well as the BriefCase device).	
Validation Process and Justification: Methods used are		Validation Process and Justification: Link to Methods		Validation Process and Justification: Link to Methods	

described in the 510(k) summary .	Description	Description AiDoc develops software within a Design Control process that is aligned with FDA 21 CFR Part 820 and ISO 62304 Medical device software - Software life cycle processes.
---	-----------------------------	--

Resources
<ul style="list-style-type: none">• Evaluation References: Please reference our clinical compendium.• Clinical Trial: N/A• Peer Reviewed Publication(s): Aidoc's clinical compendium with 100+ Peer-reviewed publications or abstract/conference presentations are available at the following link.• Reimbursement Status: N/A• Patient consent or disclosure required or suggested: N/A

References

Note: The mention or sharing of any examples, products, organizations, or individuals does not indicate any endorsement of those examples, products, organizations, or individuals by the Coalition for Health AI (CHAI). Any examples provided here are still under review for alignment with existing standards and instructions. We welcome feedback and stress-testing of the tool in draft form.

The information provided in this document is for general informational purposes only and does not constitute legal advice. It is not intended to create, and receipt or review of it does not establish, an attorney-client relationship.

This document should not be relied upon as a substitute for consulting with qualified legal or compliance professionals. Organizations and individuals are encouraged to seek advice specific to their unique circumstances to ensure adherence to applicable laws, regulations, and standards.

For instructions, references, contributors, and disclaimers please refer to the full documentation located at [www.chai.org](#).

Copyright (c) 2024 Coalition for Health AI, Inc.