









ISO 17025 CALIBRATION CERTIFICATE

Certificate Number: MNT-2782

This Calibration Certificate shall not be reproduced, except in its full entirety, without the written permission of Sensor Calibrations.

Customer

Monnit Corporation 3400 South, West Temple South Salt Lake, Utah 84115

Device

Manufacturer: Monnit Corp Model: MNS2-9-W2-HU-RH-L03

Identifier/Serial: 688479 Condition: Functional

Calibration Summary

Calibration Date: 11/16/2023 Next Calibration Date*: 6/28/2024
Calibration Procedure: 3 Lab Conditions: 39.64% RH @ 21.10°C

Result: PASS

Calibration Data

Set Point	Reference	As Found	Adjustment	As Left	Tolerance	Uncertainty†	Status
80% RH	82.4%	78.7%	N/A	78.7%	5.0%	2.3%	PASS
50% RH	50.2%	46.7%	N/A	46.7%	5.0%	2.3%	PASS
20% RH	20.0%	19.3%	N/A	19.3%	5.0%	2.1%	PASS
20°C	20.1°C	20.0°C	N/A	20.0°C	1.0°C	0.28°C	PASS
30°C	29.9°C	29.7°C	N/A	29.7°C	1.0°C	0.40°C	PASS
40°C	39.5°C	39.4°C	N/A	39.4°C	1.0°C	0.45°C	PASS

[†] Uncertainty data presented above is Total Expanded Uncertainty (TEU). The coverage factor of K = 2 results in a coverage probability of approximately 95%.

Reference Standards

Asset	Description	Cal-Date	Due-Date
S000113	Rotronic HygroClip2 Humidity Reference	10/31/2022	11/30/2023

NIST Traceability and Disclaimers

All reference standards used in this calibration are traceable to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) or an equivalent National Measurement Institute that is a signatory to CIPM MRA.

The decision rule is Simple Acceptance (SA) with a Test Uncertainty Ratio (TUR) greater than or equal to 1:1.

Details of our scope of accreditation are available at www.A2LA.org. The parameters of this calibration are covered under Sensor Calibrations' current scope of accreditation.

The results in this certificate apply only to the calibration of the device identified above.

*The listed Next Calibration Date is provided by the customer and does not imply continuing conformance during the resultant calibration interval between the Calibration Date and the Next Calibration Date.

Calibration Technician
Colton Robinson

Kelly 5 Lewis

Authorizer

Kelly S Lewis

Date Report Issued: 11/28/2023

Certified at Sensor Calibrations LLC