





Report of Human Cell Line Authentication



Report of Human Cell Line Authentication

Delivery Date: Nov 18, 2022

Analysis Date: Nov 23, 2022

I . Sample

Sample type: cell, labeled as BCPAP, was received on Nov 18, 2022.

II . Method and Procedure

1. DNA was extracted by chelex 100 reagent.
2. DNA is amplified with STR Multi-amplification Kit (Goldeneye™20A ID System);
3. PCR products are assayed with ABI 3730xl DNA Analyzer (Applied Biosystems®).
4. Data were analyzed using GeneMapper3.2 software and then compared with the ATCC and DSMZ databases for reference matching.

III. Results

1. The STR profiles of the negative and positive control match expectations.
2. The STR profiles of the cell line are in the attached tables and figure.
3. The results analysis:
 - 1) No cross-contamination of other human cell lines is found.
 - 2) The submitted profile has an 97.2%% match with the following DSMZ human cell line(s) in the DSMZ STR database (17 loci plus Amelogenin): B-CPAP.

Notes:

1. Based on the ANSI Standard, cell lines with $\geq 80\%$ match are considered to be related; i.e. derived from a common ancestry. Cell lines with between a 55% to 80% match require further profiling for authentication of relatedness.
2. This data and analysis are for research use only.

Table 1: STR profiles of the cell line

Cell line BCPAP (Fig. BCPAP)		
Marker	Allele 1	Allele 2
D19S433	13.2	15
D5S818	10	11
D21S11	30	31.2
D18S51	13	17
D6S1043	12	19
D3S1358	16	17
D13S317	12	12
D7S820	10	10
D16S539	11	12
CSF1PO	13	13
Penta D	10	11
AMEL	X	X
vWA	14	17
D8S1179	12	13
TPOX	8	11
Penta E	5	12
TH01	6	9.3
D12S391	18	23
D2S1338	18	18
FGA	20	23

Figure: STR profiles of BCPAP cell line

