



National Accreditation Board for
Testing and Calibration Laboratories

CERTIFICATE OF ACCREDITATION

**WORLDWIDE OILFIELD MACHINE PVT. LTD. -
CALIBRATION LABORATORY**

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

in the field of

CALIBRATION

Certificate Number: CC-3098

Issue Date: 16/01/2023

Valid Until:

15/01/2025

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.
(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Identity : WORLDWIDE OILFIELD MACHINE PVT. LTD.

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

1 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
Permanent Facility					
1	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Bore gauges - Only Transmission Error	Using Digimatic Plunger Dial Gauge, Dial Calibration Tester by Comparison Method	Upto 2.0 mm	0.0048mm
2	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Caliper (Vernier/ Dial/ Digimatic/ Ball type) L.C. 0.01 mm & coarser	Using Caliper Checker (Slip Gauge and Comparator Stand for Depth) by Comparison Method	0 to 609.6 mm	0.017mm
3	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Depth Gauge (Vernier/ Dial) L.C: 0.02 mm & Coarser	Using Length Bar, Caliper Checker & Granite Surface Plate by Comparison Method	0 to 609.6 mm	0.023mm
4	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	External Micrometer (L.C: 0.001 mm & coarser)	Using `0' Grade Slip Gauge / Length Bar by Comparison Method	0 to 304.8 mm	0.0040mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

2 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
5	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Feeler gauge	Using Comparator Stand with Electronic Probe by Comparison Method	0.03 mm to 1.0 mm	0.0025mm
6	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Height gauge (Vernier/ Digimatic/ Dial) L.C : 0.01 mm & Coarser	Using Length Bar, Caliper Checker , Granite Surface Plate by Comparison Method	0 to 609.6 mm	0.016mm
7	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Inside Dial Caliper L.C : 0.01mm & Coarser	Using ULM by Comparison Method	10 mm to 155 mm	0.0071mm
8	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometer (Extension Rod)	Using ULM, Length Bar by Comparison Method	12.7 mm to 304.8 mm	0.015mm
9	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Internal Micrometre (Micrometre Head) L.C: 0.01 mm & Coarser	Using ULM, Length Bar by Comparison Method	0 to 375 mm	0.015mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

3 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
10	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Type Dial Gauge L.C: 0.001 mm	Using Dial Calibration Tester with Electronic Probe & DRO by Comparison Method	0 to 0.14 mm	0.003mm
11	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Lever Type Dial Gauge L.C: 0.01 mm	Using Dial Calibration Tester with Electronic Probe & DRO by Comparison Method	0 to 1 mm	0.0066mm
12	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Measuring Pin Set	Using ULM by Comparison Method	0.45 mm to 15.24 mm	0.0013mm
13	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Micrometer Setting Standard / Height Setting Master	Comparator Stand with Electronic Probe, Slip Gauge '0' Grade / Length Bar by Comparison Method	25.0 mm to 304.8 mm	0.0066mm
14	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plain Plug Gauge	Using ULM & OD master by Comparison Method	1 mm to 203.2 mm	0.002mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

4 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
15	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Type Dial Gauge L.C: 0.01 mm & Coarser	Using Dial Calibration Tester with Electronic Probe & DRO by Comparison Method	0 to 10 mm	0.0066mm
16	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Plunger Type Dial Gauge L.C:0.001 mm & coarser	Using Dial Calibration Tester with Electronic Probe & DRO by Comparison Method	0 to 3.81 mm	0.0032mm
17	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Ring Gauge	Using ULM, Master Setting Ring Gauge by Comparison Method	100 mm to 300 mm	0.0036mm
18	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Ring Gauge	Using ULM, Master Setting Ring Gauge by Comparison Method	3 mm to 100 mm	0.0017mm
19	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauges - Major and Pitch Diameter	Using ULM, OD Master, Thread Measuring Wires by Comparison Method	100 mm to 200 mm	0.0025mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

5 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)($\pm$)
20	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauges - Major and Pitch Diameter	Using ULM, OD Master, Thread Measuring Wires by Comparison Method	200 mm to 300 mm	0.0031mm
21	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Plug Gauges - Major and Pitch Diameter	Using ULM, OD Master, Thread Measuring Wires by Comparison Method	3 mm to 100 mm	0.0018mm
22	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauges - Minor and Pitch Diameter	Using ULM, Master Setting Ring Gauge by Comparison Method	100 mm to 200 mm	0.0033mm
23	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauges - Minor and Pitch Diameter	Using ULM, Master Setting Ring Gauge by Comparison Method	200 mm to 300 mm	0.0044mm
24	MECHANICAL-DIMENSION (BASIC MEASURING INSTRUMENT, GAUGE ETC.)	Thread Ring Gauges - Minor and Pitch Diameter	Using ULM, Master Setting Ring Gauge by Comparison Method	5 mm to 100 mm	0.0017mm



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PVT. LTD. -CALIBRATION LABORATORY, GAT NO.-778, AT & POST VELU, TALUKA BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

CC-3098

Page No

6 of 6

Validity

16/01/2023 to 15/01/2025

Last Amended on

-

S.No	Discipline / Group	Measurand or Reference Material/Type of instrument or material to be calibrated or measured / Quantity Measured /Instrument	Calibration or Measurement Method or procedure	Measurement range and additional parameters where applicable(Range and Frequency)	* Calibration and Measurement Capability(CMC)(±)
25	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analogue Pressure Gauges (Hydraulic)	Using Hydraulic Ultra High Pressure Test Pump and Digital Pressure Gauge by Comparison Method	0 bar to 2760 bar	1.25bar
26	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analogue Pressure Gauges (Hydraulic)	Using Hydraulic Ultra High Pressure Test Pump and Digital Pressure Gauge by Comparison Method	0 bar to 1380 bar	1.11bar
27	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analogue Pressure Gauges (Hydraulic)	Using Hydraulic Ultra High Pressure Test Pump and Digital Pressure Gauge by Comparison Method	0 bar to 345 bar	0.15bar
28	MECHANICAL-PRESSURE INDICATING DEVICES	Digital / Analogue Pressure Gauges (Hydraulic)	Using Hydraulic Ultra High Pressure Test Pump and Digital Pressure Gauge by Comparison Method	0 bar to 69 bar	0.04bar

* CMCs represent expanded uncertainties expressed at approximately the 95% level of confidence, using a coverage factor of $k = 2$.



**National Accreditation Board for
Testing and Calibration Laboratories**

CERTIFICATE OF ACCREDITATION

**WORLDWIDE OILFIELD MACHINE PRIVATE LIMITED -
NDT LABORATORY**

has been assessed and accredited in accordance with the standard

ISO/IEC 17025:2017

**"General Requirements for the Competence of Testing &
Calibration Laboratories"**

for its facilities at

GAT NO. 778 AT POST - VELU , TAL. BHOR, PUNE, MAHARASHTRA, INDIA

in the field of

TESTING

Certificate Number: TC-9546

Issue Date: 24/05/2021

Valid Until:

23/05/2023

This certificate remains valid for the Scope of Accreditation as specified in the annexure subject to continued satisfactory compliance to the above standard & the relevant requirements of NABL.
(To see the scope of accreditation of this laboratory, you may also visit NABL website www.nabl-india.org)

Name of Legal Identity : WORLDWIDE OILFIELD MACHINE PRIVATE LIMITED

Signed for and on behalf of NABL



N. Venkateswaran
Chief Executive Officer



National Accreditation Board for Testing and Calibration Laboratories

SCOPE OF ACCREDITATION

Laboratory Name :

WORLDWIDE OILFIELD MACHINE PRIVATE LIMITED - NDT LABORATORY, GAT NO. 778 AT
POST - VELU , TAL. BHOR, PUNE, MAHARASHTRA, INDIA

Accreditation Standard

ISO/IEC 17025:2017

Certificate Number

TC-9546

Page No

1 of 1

Validity

24/05/2021 to 23/05/2023

Last Amended on

-

S.No	Discipline / Group	Materials or Products tested	Component, parameter or characteristic tested / Specific Test Performed / Tests or type of tests performed	Test Method Specification against which tests are performed and / or the techniques / equipment used
Permanent Facility				
1	MECHANICAL-MECHANICAL PROPERTIES OF METALS	Metallic material	Brinell hardness testing-HBW	ASTM E10
2	MECHANICAL-MECHANICAL PROPERTIES OF METALS	Metallic material	Rockwell Hardness Testing-HRC	ASTM E18
3	MECHANICAL-MECHANICAL PROPERTIES OF METALS	Metallic materials	Rockwell hardness testing - HRBW	ASTM E18
4	NON-DESTRUCTIVE-METALS & ALLOYS	API 20 E Bolting Material, Round Bars & Forgings	Ultrasonic Testing, Volumetric NDE , technic - pulse echo	ASME Section V, ASTM A388
5	NON-DESTRUCTIVE-METALS & ALLOYS	API 20 E Bolting Material, Ferrous & non ferrous metals & alloys	Surface crack detection , Liquid penetrant testing , technic - solvent removable, visible	ASME Section V, ASTM E165
6	NON-DESTRUCTIVE-METALS & ALLOYS	Ferromagnetic materials & API 20E bolting material	Magnetic Particle testing wet Fluorescent Technic - circular and longitudinal magnetisation	ASME Section V, ASTM E709