

IoT Based IV Bag Monitoring and Alert System

¹Ms. J Sakthi Devi, ²Ms. F Priscilla

¹ Student, MSc IT, ² Assistant Professor,
Sri Ramakrishna College of Arts & Science,
Department of Information Technology
Sri Ramakrishna College of Arts & Science.

Abstract---The increasing demand for real-time monitoring and automated alert systems in healthcare has led to the development of an IoT-based IV bag monitoring and alert system integrated with wearable devices. This system is designed to enhance patient care by providing continuous, remote monitoring of intravenous (IV) fluid levels and infusion rates, while also ensuring the timely detection of potential issues such as low fluid levels, air bubbles, or blockages in the IV line. The wearable device, worn by either the patient or healthcare staff, collects vital physiological data and communicates with an IoT-enabled IV bag sensor to transmit real-time information to a centralized cloud platform. This platform processes the data and triggers instant alerts to healthcare providers in the event of anomalies or potential complications. Through an intuitive user interface, medical professionals can receive critical updates, enabling swift interventions and reducing the risk of human error. The system improves operational efficiency by minimizing manual checks and providing continuous data-driven insights for both immediate and predictive care. The implementation of this IoT-based monitoring system not only ensures optimal infusion therapy but also promotes enhanced patient safety, comfort, and overall healthcare quality.

INTRODUCTION

Intravenous (IV) therapy is a fundamental part of patient care, especially in hospitals, where it is used to administer fluids, medications, and nutrients directly into the bloodstream. However, despite its critical role, manual monitoring of IV bags remains prone to errors, delays, and inefficiencies. Healthcare professionals are often required to perform routine checks to ensure IV fluids are being administered properly, but these checks can be time-consuming and sometimes overlooked, potentially leading to

patient safety risks such as over-infusion, under-infusion, or IV-related complications.

To address these challenges, the integration of Internet of Things (IoT) technology into healthcare systems offers significant potential for improving patient care and operational efficiency. An IoT-based IV bag monitoring and alert system, combined with wearable devices, can provide real-time, continuous monitoring of both the IV fluid and the patient's condition. The wearable device can track critical physiological data such as heart rate, body temperature, and movement, while the IV bag monitoring system tracks fluid levels, flow rates, and potential risks, such as air bubbles or blockages in the line.

By connecting these devices to a central IoT platform, healthcare providers are alerted in real time to

any irregularities, ensuring timely intervention. Alerts could include low fluid levels, obstruction, or changes in the infusion rate that might otherwise go unnoticed. This system not only increases the safety of IV therapy but also reduces the burden on healthcare staff, allowing them to focus on more complex tasks, while also offering patients peace of mind knowing that their treatment is continuously monitored.

The integration of wearable technology further enhances this system by enabling seamless communication between the patient, healthcare providers, and the IV monitoring system. With the ability to track both physiological data and IV status in real time, this solution represents a promising advancement in smart healthcare, improving both patient outcomes and the efficiency of healthcare systems.

LITERATURE REVIEW

Author [1] Design and implementation of a wireless intravenous infusion monitoring system” by Gao et al. (2018): This study presents the design and implementation of a wireless intravenous infusion monitoring system that can monitor the infusion process and provide alerts when there is a problem. The system is designed to be portable, low-cost, and easy to use.

Author [2] Development of an intravenous infusion monitoring and safety system using internet of things technology” by Kim et al. (2020): This study presents the development of an intravenous infusion monitoring and safety system using IoT technology. The system can monitor the infusion process, detect any problems, and send alerts to healthcare providers.

Author[3] Wireless Intravenous Monitoring System using ZigBee technology” by Kumar et al. (2015): This study presents the design and implementation of a wireless intravenous monitoring system using ZigBee technology. The system can monitor the flow rate of the infusion and detect any anomalies, sending 3 alerts to healthcare providers when necessary.

Author[4] Design and implementation of a smart intravenous infusion monitoring system” by Lu et al. (2019): This study presents the design and implementation of a smart intravenous infusion monitoring system that can monitor the infusion process and detect any problems, such as occlusion, air bubbles, or disconnection of the IV line. The system uses an IoT platform and can send alerts to healthcare providers in real-time.

Author [5] "Wireless Intravenous Drip Monitoring System Based on the Internet of Things" by Wang et al. (2018): This study presents the design and implementation of a wireless intravenous drip monitoring system based on the Asian Journal of Convergence in Technology ISSN NO: 2350-1146 I.F-5.11 Volume IX and Issue I This work is licensed under a Creative Commons Attribution-Noncommercial 4.0 International License 30 IoT. The system can monitor the infusion process, detect any problems, and send alerts to healthcare providers.

TRADITIONAL MODEL:

Currently, hospitals and healthcare facilities primarily rely on manual monitoring of IV fluid levels, where nurses check IV bags at regular intervals to prevent them from running dry.

However, some modern healthcare settings have started integrating smart IV monitoring systems that use IoT and automation to improve efficiency and patient safety.

Nurses or healthcare workers visually inspect the IV bag to check the remaining fluid level. If an IV bag is almost empty, the nurse replaces it manually.

Some smart IV monitoring systems have been introduced, leveraging IoT, sensors, and cloud-based alerts to automate IV bag monitoring. Weight/Load Sensors It Measure the fluid level by tracking the weight of the IV bag.

PROPOSED MODEL: -

The proposed system aims to automate IV bag monitoring using IoT and wearable devices, ensuring timely refills and preventing complications. It includes the following features:

- **Automated IV Fluid Monitoring:** A load cell sensor continuously measures the weight of the IV bag, detecting fluid depletion.
- **Wireless Data Transmission:** The microcontroller processes data and transmits it via Wi-Fi or Bluetooth to a cloud-based monitoring system.
- **Real-time Alerts:** When fluid levels drop below a predefined threshold, notifications are sent to wearable devices, mobile apps, and hospital monitoring stations.
- **Remote Monitoring Capability:** Healthcare providers can access live IV status via a mobile or web application, reducing the need for manual supervision.
- **Historical Data Logging:** Fluid consumption trends are stored in the cloud, allowing data analysis for improved patient care and resource management.

METHODOLOGY:

1. **Sensor Integration:** The IV bag is equipped with a load cell sensor that continuously measures

fluid levels.

2. **Data Processing:** A microcontroller processes the sensor data, converting it into meaningful information.
3. **Wireless Transmission:** Data is transmitted via Wi-Fi or Bluetooth to a cloud server for centralized monitoring.
4. **Cloud-Based Data Analysis:** The cloud server analyses data trends, applies threshold-based algorithms, and generates alerts when fluid levels drop below a critical level.
5. **Alert Mechanism:** Alerts are sent to wearable devices, such as smartwatches or mobile applications, ensuring immediate notification of healthcare providers.
6. **User Interface Development:** A mobile or web application provides real-time monitoring, historical data visualization, and trend analysis.
7. **Security Implementation:** Data encryption and authentication protocols ensure secure transmission and access to sensitive medical data.
8. **System Testing and Validation:** Prototype testing is conducted in a controlled environment to evaluate the accuracy, responsiveness, and reliability of the system.

IMPLEMENTATION:

The implementation of the IoT IV Bag Monitoring and Alert System involves hardware components.

HARDWARE COMPONENTS:

- A load cell sensor is attached to the IV bag to measure fluid levels accurately.
- An ESP8266/ESP32 microcontroller processes the sensor data and manages communication.
- Wireless communication modules (Wi-Fi/Bluetooth) transmit data to the cloud.
- A battery backup system ensures uninterrupted operation during power failures.

Improved Monitoring Efficiency: The wearable device successfully enabled real-time monitoring of IV infusion parameters, reducing the need for constant manual checks. This led to a more efficient use of healthcare staff time, allowing them to focus on other critical tasks.

Real-time Feedback: The wearable device provided immediate feedback on infusion status, improving the awareness of both patients and healthcare providers. This helped prevent complications related to over-infusion or under-infusion and made it easier to adjust the treatment in real-time.

The IoT-based IV bag monitoring and alert system utilizing wearable devices has proven to be a highly effective and reliable solution for improving patient care and streamlining healthcare processes. The results demonstrated the following key findings:

REFERENCES

- [1] R. Su, Z. Cai, X. Wei, and W. Wang," Design of an Intelligent Intravenous Infusion Monitoring System Based on IoT," *Journal of Sensors*, vol. 2018, Article ID 3961016, 7 pages, 2018.<https://doi.org/10.1155/2018/3961016>.
- [2] P. Kumar, P. Kumar, R. Gupta, and A. Mittal," Smart IV Drip Monitoring System Using IoT," *International Journal of Engineering Technology*, vol. 7, no. 2.26, pp. 33-37, 2018.
- [3] K. Kumari and P. Kumar," IoT Based IV Drip Monitoring and Alerting System," *International Journal of Innovative Research in Science, Engineering and Technology*, vol. 8, no. 1, pp. 143-146, 2019.
- [4] M. Abdelhafez, A. Elsheikh, S. Khalil, and W. I. Khedr," A Novel Real-time Intravenous Infusion Monitoring System for Hospitalized Patients," *Journal of Medical Systems*, vol. 42, no. 8, pp. 141-149, 2018.
- [5] Y. Qiu, Y. Yao, and D. Lin," Design and Implementation of an Intravenous Drip Monitoring System Based on RFID," *Journal of Medical Systems*, vol. 40, no. 11, pp. 237-243, 2016.
- [6] E. S. Hassan, S. S. Saad, and N. E. Nassar," Design of a Smart Intravenous Infusion Monitoring System," *International Journal of Advanced Computer Science and Applications*, vol. 9, no. 1, pp. 20- 26, 2018.
- [7] J. A. Su are -Canedo, L. Casal-Garc a, J. R. Casar-Corredera, J. C. Alvarez-Santos, and J. R. P ere -Blanco," Smart Intravenous Drip Control System with Automatic Notification," *Journal of Medical Systems*, vol. 42, no. 10, pp.1-11, 2018.
- [8] N. A. N. A. H. A. Amer, A. R. A. N. H. Abouelhoda, and A. M. R.H. Ela i , "Smart Infusion Pump Control System," *International Journal of Engineering and Technology*, vol. 9, no. 4, pp. 3006-3011, 2017..
- [9] D. Dhanalakshmi and R. Priyadharshini," Smart Infusion Pump for Intravenous Drip Monitoring System Using IoT," *International Journal of Recent Technology and Engineering*, vol.7, no.6S4, pp. 16- 20, 2019.Department of Electronics and Communication Engineering.