

20250819_CS_5C_sort Report_001




Sort Details

Date	Start Time	Instrument Configuration	Sort Threshold Count	Group Name
August 19, 2025	12:59:03 PM	5-Laser-V16-B14-R8-YG10-UV16	10,366,747	Group_001
Operator	Stop Time	Instrument Serial Number	E-Abort Count	Tube Name
Admin	15:08:03 PM	S0172	290,457	Lib1 no pen strep

Sort Parameters

Nozzle Size and S/N	Sort Mode	Drop Delay Value	Amplitude	Break-Off - Setpoint	Drop Interval - Setpoint	Plate Voltage (V)	Center Stream Optimization	Temperature Control
100 µm K51002	Purity	8623	11536	218	20	3,000 V	4th drop: 2, 3rd: 4, 2nd: 7, 1st: 15	37
Nozzle Setting	Sort Type	Drop Drive Frequency	Sort Monitoring	Break-Off - Actual	Drop Interval - Actual	Drop Charge	Sheath Pressure	Mixing
Default	Tube	24000	On	220	20	L:-31, R:39, MR:101	17.6 psi	Medium

Tubes

	LEFT	RIGHT	MIDDLE RIGHT
	Population Name P4 Total Count 512618 	Population Name P5 Total Count 84142 	Population Name P7 Total Count 32106 
1	Starting Volume 1000.00 µl Average Sort Rate 73 e/s Average Abort Rate 16 e/s Average Efficiency 83% Sort Threshold Count 10366747 % Total 5.9931 Sorted Volume 2660.62 µl Count 512618 Abort Count 108676 Quick Stop None Waste Tank Full False	Starting Volume 1000.00 µl Average Sort Rate 12 e/s Average Abort Rate 9 e/s Average Efficiency 58% Sort Threshold Count 10366747 % Total 1.404 Sorted Volume 440.12 µl Count 84142 Abort Count 61407 Quick Stop None Waste Tank Full False	Starting Volume 1000.00 µl Average Sort Rate 5 e/s Average Abort Rate 3 e/s Average Efficiency 64% Sort Threshold Count 10366747 % Total 0.4837 Sorted Volume 167.88 µl Count 32106 Abort Count 18039 Quick Stop None Waste Tank Full False

Comments

Submit Comment