

20250819_CS_5C_sort Report_002

Sort Details

Date	Start Time	Instrument Configuration	Sort Threshold Count	Group Name
August 19, 2025	15:39:34 PM	5-Laser-V16-B14-R8-YG10-UV16	3,368,804	Group_001
Operator	Stop Time	Instrument Serial Number	E-Abort Count	Tube Name
Admin	16:49:07 PM	S0172	93,329	Lib1 with penstrep


Sort Parameters

Nozzle Size and S/N	Sort Mode	Drop Delay Value	Amplitude	Break-Off - Setpoint	Drop Interval - Setpoint	Plate Voltage (V)	Center Stream Optimization	Temperature Control
100 µm K51002	Purity	8623	11581	218	20	3,000 V	4th drop: 2, 3rd: 4, 2nd: 7, 1st: 15	37
Nozzle Setting	Sort Type	Drop Drive Frequency	Sort Monitoring	Break-Off - Actual	Drop Interval - Actual	Drop Charge	Sheath Pressure	Mixing
Default	Tube	24000	On	220	20	L:-31, R:39	17.6 psi	Medium

Tubes


LEFT

Population Name
P4
Total Count
139365



RIGHT

Population Name
P5
Total Count
47667



1	Starting Volume	Sorted Volume	Starting Volume	Sorted Volume
	1000.00 µl	723.02 µl	1000.00 µl	248.43 µl
	Average Sort Rate	Count	Average Sort Rate	Count
	43 e/s	139365	15 e/s	47667
	Average Abort Rate	Abort Count	Average Abort Rate	Abort Count
	13 e/s	41456	8 e/s	27491
	Average Efficiency	Quick Stop	Average Efficiency	Quick Stop
	77%	None	63%	None
	Sort Threshold Count	Waste Tank Full	Sort Threshold Count	Waste Tank Full
	3368804	False	3368804	False
	% Total		% Total	
	5.3675		2.231	

Comments

Submit Comment