

20250819_CS_5C_sort Report_003

Sort Details

Date	Start Time	Instrument Configuration	Sort Threshold Count	Group Name
August 19, 2025	16:50:08 PM	5-Laser-V16-B14-R8-YG10-UV16	5,446,216	Group_001
Operator	Stop Time	Instrument Serial Number	E-Abort Count	Tube Name
Admin	18:08:06 PM	S0172	151,163	Lib1 with penstrep


Sort Parameters

Nozzle Size and S/N	Sort Mode	Drop Delay Value	Amplitude	Break-Off - Setpoint	Drop Interval - Setpoint	Plate Voltage (V)	Center Stream Optimization	Temperature Control
100 µm K51002	Purity	8623	11528	218	20	3,000 V	4th drop: 2, 3rd: 4, 2nd: 7, 1st: 15	37
Nozzle Setting	Sort Type	Drop Drive Frequency	Sort Monitoring	Break-Off - Actual	Drop Interval - Actual	Drop Charge	Sheath Pressure	Mixing
Default	Tube	24000	Off	220	20	L:-31, R:39	17.6 psi	Medium

Tubes


LEFT

Population Name
P4
Total Count
302845



RIGHT

Population Name
P5
Total Count
51319



1	Starting Volume	Sorted Volume	Starting Volume	Sorted Volume
	1000.00 µl	1572.42 µl	1000.00 µl	267.68 µl
	Average Sort Rate	Count	Average Sort Rate	Count
	69 e/s	302845	12 e/s	51319
	Average Abort Rate	Abort Count	Average Abort Rate	Abort Count
	19 e/s	82741	8 e/s	35079
	Average Efficiency	Quick Stop	Average Efficiency	Quick Stop
	79%	None	59%	None
	Sort Threshold Count	Waste Tank Full	Sort Threshold Count	Waste Tank Full
	5446216	False	5446216	False
	% Total		% Total	
	7.0799		1.5864	

Comments

Submit Comment