

Chapter 1

Conclusions

In this work an light-sheet fluorescence microscope was designed built and optimised for organismal and cellular imaging.

1.1 Microscope design and construction

The microscope presented in ?? was able to record imaging volumes at a rate of $512\text{ }\mu\text{m} \times 512\text{ }\mu\text{m} \times 100\text{ }\mu\text{m}$ volumes per second, at maximum running speed. It featured four laser channels spanning the visible spectrum and suitable for standard fluorescent dyes. The microscope generated light-sheets digitally by sweeping a galvanometric mirror converting a standard Gaussian beam into a Gaussian sheet. The optical elements on the illumination board consisted of a beam splitter for creating two beams and an optical Spatial Light Modulator (SLM) path for generating exotic illumination or two beams directly, steering mirrors and relay lenses. A XYZ stage was used to mount and position samples with sub- μm precision. The imaging plane of the detection objective was matched to the scanning light-sheet by using a fast piezo electric objective actuator. An additional objective lens relay was positioned in front of the detection camera to increase the total magnification from $25\times$ to $31.25\times$ or $62.5\times$ when using the $1.25\times$ and $2.5\times$ objective lenses respectively. This provided a variable field of view for imaging organisms and single cells, with the largest magnification being oversampled compared to Nyquist's limit.

1.2 Homographically generated light-sheets

In Chapter ?? the method by which the generation of voltage waveforms was improved upon. Standard methods of generation rely on three positional coordinates being record for registering the illumination volume to the imaging volume. To use the four available coordinates additional non-linear mathematics is needed to project affine input coordinates onto the resultant non-affine illumination space. The non-linearity observed (figure) arises from imperfections in the scanning optics, this was characterised for the scan lens which converts beam angle into beam position at the sample volume. When comparing the 3 point and 4 point correction it was found that there was an average 42 % increase in recovered signal from fluorescent beads.

1.3 Frame localisation optical projection tomography

Chapter ?? then used the mathematics presented in Chapter ?? to address the issue of computer-tomography reconstructions in systems prone to mechanical error; in this case for optical projection tomography. The ubiquitous Radon transform is typically used to reconstruct projected data sets but take no account of motions beyond rotation about the axis normal to the plane projected. This assumption is valid for stable systems, but leads to reconstruction error for systems with limited available angles as well as systems with mechanical jitter and systematic drift. Provided there are 6 or more fiducial markers common to each adjacent image pair from an Optical Projection Tomography (OPT) image set, each point can be triangulated and the resultant rotation and translation matrices between each image pair can be computed. Each image from the set can then be computational back-projected and aligned using the recovered matrices. The resultant volume is then filtered to provide a faithful reconstruction of the volume that is robust to systematic mechanical errors as well as jitter.

1.4 Confocal slit-scanning microscopy

The illumination and imaging volume registration was also critical for Chapter ?? where synchronisation of the rolling shutter on the camera and the sweeping of the beam had to be well matched. By synchronising the rolling shutter and the beam sweeping images can be recorded with better optical sectioning and higher

contrast as out-of-focus light is rejected by the virtual pinhole in the direction of the rolling shutter. This was implemented on the light-sheet microscope and was characterised to show the maximal Signal to Noise ratio (SNR) was when the shutter width and the beam waist were comparable in size. The firmware on the camera was then upgraded along with the beam creation optics such that two shutters would roll synchronously and two beams would be generated to match them. This then increased the achievable frame-rate of the slit-scanning camera from 50 Hz to 100 Hz, the maximum of the camera. This was in preparation for the imaging capabilities needed for tracking virus particles as in Chapter ?? . The effect of slit-scanning was also simulated for light-sheet systems as well as wide-field systems. This included structured illumination systems where it was shown that a marked rate of imaging and contrast improvement could be obtained.

1.5 Open-hardware sample mounting

In preparation for the two chapters addressing biological questions a sample chamber was designed and presented in Chapter ?? . The chamber was a Three Dimensional (3D) printed design capable of mounting cells and organisms, with fixed cells and live organisms being shown in this work. 3D printed materials were used for the rapid prototyping capability as well as the advantage of producing multiple copies quickly when compared to metal. It is possible to sterilise 3D though the more aggressive methods were damaging to the printed chambers.

The key novelty of the chamber was the use a shelf for sample mounting. This allowed for the mounting of flat glass coverslips, which are viable for use in common cell and organism mounting procedures. It also allowed for the use of a large opening angle detection objective lens mounted at 45° in conjunction with a high NA illumination objective lens. Enabling the use of these two objectives allowed for the imaging of organisms as well as sub-cellular activities in cell culture. The chamber include addition design features such as: chamfered edges in the medium containing section to act as hard stops if protecting the objective lenses and minimise the use of media; a positing camera and window from with a couple infrared and white light source for transmitted light imaging; a side window for positing samples by eye; a drip-tray for instrument safety against spillages; module for inserting heating elements and magnets to be used for securing samples and cleared through-holes for bolting the chamber to a metric breadboard.

1.6 Particle tracking for viral egress

With such a chamber now available, imaging *in vivo* sub-cellular processes was now possible, see Chapter ?? . To then monitor viral egress, the light-sheet microscope required modifying such that it could axially localise single particles within the light-sheet. From there the light-sheet would be repositioned and the process iterated for each recorded frame giving a sub-diffraction limited axial position over the viable 100 μm . As such, a weakly cylindrical lens was inserted into the imaging path and positioned to provide a suitable amount of astigmatism for the available pixels per (4×4) Point Spread Function (PSF) at 62.5 (25×2.5 magnification. It was shown that the calibration for a recorded astigmatic PSF was suitably linear for the purposes of calibrated tracking when using template correlation or template covariance.

Dynamic blur for an astigmatic PSF was then considered. By using a Gaussian model of an astigmatic PSF it was shown that, using typical parameters as found for the light-sheet microscope here, there is an average 25 % error axial localisation. The magnitude of this error could potentially explain why single particles do escape tracking before they bleach as demonstrated by Spille *et. al* [1].

Cells infected with Herpes Simplex Virus (HSV)-1 were then fixed and mounted on glass cover slips for imaging. Three viral proteins were labelled to demonstrate the feasible of tracking single virus particles in a live cell.

1.7 Light-sheet microscopy and magnetic tweezers

Finally, in Chapter ?? a magnetic tweezer system was developed and used to remotely measure the viscoelastic properties of a developing zebrafish embryo. A simple mechanical model of the response of the embryo was made, consisting of dash-pots and springs, and was fitted to the trajectory of a tracked bead from push-pull experiments. Viscoelastic properties were then extracted from the model fitting and then monitored over 1 h to 2 h. This was the first demonstration of a direct link between cell behaviours and embryo morphogenesis arising from changing mechanical properties. The blastula tissue was shown to have an intrinsic viscoelastic response. The stiffening blastula, composed of radially anisotropic cells, may drive blastula thinning and yolk bulging in the next morphogenetic movement of the zebrafish embryo. Interestingly, the viscoelastic timescale parameter, τ , remained constant across both time and embryonic mutations.

Chapter 2

Perspectives

Looking forward there are several key areas with which this body of work could be extended and improved upon.

2.1 Light-sheet microscope developments

2.1.1 Illumination

The system would benefit from having thinner light-sheets, particularly for single cellular imaging where axial resolution is more important when compared with larger organism imaging. Exotic illumination modes were explored briefly during this work and a Bessel beam was constructed using the SLM as featured in Chapter ???. As discussed Bessel beams can be used to create thinner light-sheets over a range of $\sim 40\text{ }\mu\text{m}$. Similarly, Airy beam light-sheets could be created using a cubic phase mask on the SLM. Airy Beams behave over a longer range ($500\text{ }\mu\text{m}$) compared to Bessel beams though they require deconvolution in the final step which can take a long time with large data sets. Alternatively a 2-Photon (2P) beam could also be used to create a light-sheet thinner than single photon illumination, though infrared lasers are expensive and would require new optics.

2.1.2 Imaging

During this work a plug-and-play Fluorescence Lifetime Imaging Microscopy (FLIM) enabled camera became commercially available for wide-field microscopes, combined with the presented light-sheet microscope this could provide a fast volumetric

fluorescence lifetime readout from cells to organisms. Stelzer *et. al* showed that volumetric light-sheet combined with FLIM can be useful in organismal imaging [2]. FLIM combined with the viscoelastic measurements seen in the Chapter ?? could further elucidate tensions occurring locally to the bead under motion, with a suitable fluorescent reporter [3]. Fluorescence lifetime could also be used for indicating if a virion has moved between boundaries within a cell or collected tegument proteins, though the SNR of single particle lifetime measurement may be too low.

2.2 Waveform generation

For waveform generation a 4 pt calibration was presented which gave a 42 % improvement in signal collecting efficiency, by matching the illumination and imaging volumes better than a 3 pt calibration. It is possible to create a 5 pt calibration using elastic transforms (elastic transforms are needed as the projective mathematics in Chapter ?? breaks down beyond 4 pts), which could further increase the collection efficiency. With calibrations beyond 4 pts it becomes difficult to find the corresponding point between the beam and the image. The most obvious next point to choose beyond the 4 available corners of the illumination volume, would be the centre point of the those corners at $z : 100 \mu\text{m}$, $x : 1024 \text{ px}$. Finding the focus at this z position would involve maximising the image maximum when imaging in dye (as in Chapter ??), matching the 1024th pixel with the middle of the beam would be more difficult, particularly by eye. As a result, it would best to position the 5th point calibration automatically by using a peak finding algorithm for the axial (z) focus matching, using image maxima; and in (x), by maximising the summed intensity along the requisite pixel row. In doing so, all other available pixel rows in x would be trivial to calibrate to, as would all axial positions. From this, a complete look-up table of positions could be created for the scanning beam. Such an approach would mean each imaging session would require a potentially lengthy calibration, but for the benefit of not needing to generate waveforms during imaging and having very precisely focused light-sheets.

2.3 Slit-scanning

2.3.1 Light-sheet slit scanning

Confocal slit-scanning was found to complement Bessel beam illumination by rejecting light generated from the higher orders of the beam [4]. By using the dual-beam optics as presented in Chapter ??, two Bessel beams could be created to provide 100 Hz Bessel sheet imaging for high axial resolution sub-cellular imaging.

2.3.2 Wide-field slit-scanning

The slit-scanning theory presented in Chapter ?? for non-light-sheet microscopes will need to be implemented on a real system to validate the simulations presented in Chapter ?. The National Physical Laboratory has such a Structured Illumination Microscopy (SIM) available for use, with a sufficiently fast SLM to match the speed of the rolling shutter with a scanning illumination line.

2.3.3 Slit-scanning fusion imaging

A further extension of this work could involve using two cameras and an image splitter. By rotating the additional camera at the second port of the image splitter and superimposing two orthogonal sweeping lines on the SLM, two images of the same sample would be produced with increased contrast and resolution in orthogonal directions. A joint Richardson-Lucy deconvolution [5] could then be used to fuse the two images into a homogenous high resolution and high contrast image. As an image splitter would be used, a brighter image would be needed to maintain the SNR, this could be achieved using more illumination light or increasing the overall exposure time. The method proposed would be able to achieve potential super-resolution at video rates.

2.4 Frame localisation optical projection tomography

Chapter ?? presented a theoretical framework for the reconstruction of OPT data sets that was robust to systematic mechanical drift and noise. To verify that the reconstruction works beyond simulated data, an OPT image set with 5 or more fiducial beads will need to be reconstructed. Watson *et al.* [6], have published a set of

mechanical adaptors and softwares so that such a data set could be made using a standard inverted fluorescence microscope. Given an appropriate dataset, a particle tracking algorithm for the labelling trajectories that is robust to occlusions will also be needed.

2.4.1 Fiducial-free optical projection tomography

In terms of developments to the theory presented in Chapter ??, it may be possible to produce a fiducial-free reconstruction provided the sample has sufficient visual texture. Texture here refers to variance of contrast within the image, an example would be the transmitted white light image of a developed Zebrafish. In computer vision, the fundamental, essential and homography matrices are found from image pairs with no placed fiducial markers. Instead, features that are common to the each view are used as points for triangulation. The confidence in such points is much lower than the defined manually placed fiducial marks, but, many more can be used, increasing the overall confidence in the recovered matrices. From here, the recovered matrices can be deconstructed and deployed as described in Chapter ??. This could be implemented provided a transmission Optical Projection Tomography (tOPT) dataset were available.

2.5 The sample chamber

The current sample chamber design as shown in Chapter ?? does not provide atmospheric control beyond closed-loop temperature control, which is needed for some organisms and CO₂ buffered media. To make the chamber viable for long-term live cellular imaging, a suitable lid would need to be designed and added to the chamber. The top of the sample chamber was purposefully designed flat so that such a lid could be created and attached. The intention was to use the standard chamber to position a sample and then attach the appropriate lid for long term imaging. The lid would need an entry point for gas flow and the lid would be supported by the walls of the chamber below. Gaskets would be fitted around the objectives to reduce gas loss and an O-ring would couple the lid and sample chamber to a mostly air-tight fit, though some flow would be needed to not pressurise the vessel.

2.5.1 User construction

Ideally the chamber would be printed with the option to have clear windows without the need for user intervention i.e. glueing perspex windows in manually. However, advances in the 3D printing technology would be needed before the chamber could be printed from a single piece. This would require optically clear printing which, using extrusions techniques currently, is impossible. It may be possible to 3D a design for casting a sample chamber in one piece from a curable plastic, however this method would be much more involved than the method proposed.

2.6 Viral imaging using single particle tracking

It was shown in Chapter ?? that the light-sheet microscope was capable of localising a static single particle. However, efforts are needed to demonstrate this in dynamic systems. One approach would be tracking a single particle in a known viscosity liquid as discussed in Chapter ?. Another approach could involve keeping the light-sheet static and measuring the ensemble statistics of many particles being tracked. The latter approach would, however, lead to a bias towards short tracks of particle trajectories.

2.6.1 Viral imaging in live cells

Virion movement would need to be demonstrated in live cells to show that cells survive and behave as expected in the light-sheet microscope; and that virions can be tracked when moving within a cell, at speed.

2.7 Magnetic tweezers and light-sheet microscopy

The combined technique of remote force measurement using magnets and light-sheet microscopy is now being developed into a stand-alone device in the Cambridge University Engineering Department by Fergus Riche. The device will continue to be applied to observing the mechano-biology of zebrafish as well as the characteristics of tissues. The new device is based around a flat light-sheet design, with a chamber constructed around the objectives permanently embedded with quadrupole magnetic tweezers.

2.8 Additional samples

Mammalian cells and Zebrafish were the specimens of choice for this work as they are both optically clear and biological relevant models. Human and mouse embryos were also considered as they are more biologically relevant for the study of humans and mammals. Issues with mouse embryos are that of collection yield, with the required sacrifice of a mouse to produce a mouse embryo, as well as the atmospheric conditions required to image the embryo. Human embryos are also much less abundant than Zebrafish embryos as well as requiring a very specific environment to survive. That said, given the capabilities of the chamber developed in Chapter ??, both options could be explored as both embryo types would benefit from the fast and gentle volumetric imaging provided by light-sheet microscopes.

Organoids, which are miniaturised forms of organs built or grown from organ samples, would also benefit from light-sheet imaging. The issue for imaging organoids lies with them being often very opaque to visible light. Recent advances in clearing techniques and expansion microscopy would make the imaging of large fixed organoids (including brain organoids) viable for light-sheet microscopy. Clearing techniques use chemical regents to wash away scattering material from samples, particularly highly scattering lipids, to reduce the opacity of the sample [7]. Similarly expansion microscopy washes away all but the fluorophores and a gel scaffold leaving a fluorescent ghost of a sample. Water is then added to linearly swell the sample to many times its original size for super-resolved imaging for diffraction limited microscopes [8]. Light-sheet microscopy would be particularly valuable for expanded samples as the amount of time required to image an expanded *Drosophila* brain, for example, can be on the order of months using laser scanning techniques. Both clearing and expansion techniques suffer from sample fragility and collection yield due to the harshness of their respective processes.

References

1. Spille, J.-H. *et al.* Direct Observation of Mobility State Transitions in RNA Trajectories by Sensitive Single Molecule Feedback Tracking. *Nucleic Acids Research*, 1–11 (2015).
2. Greger, K., Neetz, M. J., Reynaud, E. G. & Stelzer, E. H. K. Three-Dimensional Fluorescence Lifetime Imaging with a Single Plane Illumination Microscope Provides an Improved Signal to Noise Ratio. *Optics Express* **19**, 20743–20750 (2011).
3. Colom, A. *et al.* A Fluorescent Membrane Tension Probe. *Nature Chemistry*, 1 (2018).
4. Fahrbach, F. O., Voigt, F. F., Schmid, B., Helmchen, F. & Huisken, J. Rapid 3D light-sheet microscopy with a tunable lens. *Optics Express* **21**. bibtex: fahrbach_rapid_2013-1, 21010–21026 (2013).
5. Ingaramo, M. *et al.* Richardson–Lucy Deconvolution as a General Tool for Combining Images with Complementary Strengths. *ChemPhysChem* **15**, 794–800 (2014).
6. Watson, T. *et al.* OPTiM: Optical Projection Tomography Integrated Microscope Using Open-Source Hardware and Software. *PLOS ONE* **12**, e0180309 (11-Jul-2017).
7. Clearing and Labeling Techniques for Large-Scale Biological Tissues. *Molecules and Cells*. bibtex: _clearing_2016 (2016).

8. Chen, F., Tillberg, P. W. & Boyden, E. S. Expansion microscopy. *Science* **347**. bibtex: chen_expansion_2015, 543–548 (January 2015).