



Device: Claria MRI™ Quad CRTD DTMA1QQ

Serial Number: RPA239938H

Date of Interrogation: 20-Jul-2025 04:38:25

Patient: Bolton, Genevieve

Physician: ASSAR, MANISH 469-800-7400

Device Status (Implanted: 24-Jan-2020)

Remaining Longevity 3.9 years (20-Jul-2025)



(based on initial interrogation)

	Atrial(5076)	RV(6935M--)	LV(4798)
Pacing Impedance	380 ohms	380 ohms	1045 ohms
Defibrillation Impedance		RV=73 ohms	
Pace Polarity	Bipolar	Bipolar	
Capture Threshold	0.500 V @ 0.40 ms	0.750 V @ 0.40 ms	
Measured On	20-Jul-2025	20-Jul-2025	
Programmed Amplitude/Pulse Width	1.50 V / 0.40 ms	2.00 V / 0.40 ms	
Measured P/ R Wave	3.0 mV	3.6 mV	
Programmed Sensitivity	0.30 mV	0.30 mV	

Parameter Summary

Mode	AAI<=>DDD	Lower Rate	60 bpm	Paced AV	180 ms
Mode Switch	150 bpm	Upper Track	130 bpm	Sensed AV	150 ms
		Upper Sensor	130 bpm		

Detection

AT/AF	Monitor
VF	On
FVT	OFF
VT	OFF

Rates

>150 bpm
>182 bpm

Therapies

All Rx Off
ATP During Charging, 35J x 6
All Rx Off
All Rx Off

Enhancements On: VT Monitor, AF/Afl, Sinus Tach, Wavelet, TWave, Noise

Clinical Status

Since 19-Jun-2025

Treated

VF	0
FVT (Off)	
VT (Off)	
AT/AF(Monitor)	

Monitored

VT (162-182 bpm)	0
VT-NS (>4 beats, >182 bpm)	0
High Rate-NS	0
SVT: VT/VF Rx Withheld	0
V. Oversensing-TWave Rx Withheld	0
V. Oversensing-Noise Rx Withheld	0
AT/AF	0
Time in AT/AF	0.0 hr/day (0.0%)

Functional**Last Week**

Patient Activity	1.1 hr/day
------------------	------------

Cardiac Compass Trends (May-2024 to Jul-2025)