



Device: Amplia MRI™ Quad CRTD DTMB1QQ

Serial Number: RPE203384H


Date of Interrogation: 20-Jul-2025 02:26:02

Patient: Hart, Michael Leon

Physician: AMIN - - -

Device Status (Implanted: 21-Oct-2016)

Remaining Longevity 3 months (20-Jul-2025)



(based on initial interrogation)

	Atrial(5076)	RV(6935M)	LV
Pacing Impedance	456 ohms	475 ohms	513 ohms
Defibrillation Impedance		RV=71 ohms	
Pace Polarity	Bipolar	Bipolar	
Capture Threshold	0.500 V @ 0.40 ms	0.875 V @ 0.40 ms	
Measured On	19-Jul-2025	20-Jul-2025	
Programmed Amplitude/Pulse Width	1.50 V / 0.40 ms	2.00 V / 0.40 ms	
Measured P/ R Wave	3.8 mV	15.3 mV	
Programmed Sensitivity	0.30 mV	0.30 mV	

Parameter Summary

Mode	AAI<=>DDD	Lower Rate	60 bpm	Paced AV	180 ms
Mode Switch	171 bpm	Upper Track	130 bpm	Sensed AV	150 ms
		Upper Sensor	120 bpm		

Detection

AT/AF	Monitor
VF	On
FVT	OFF
VT	On

Rates

>171 bpm
>207 bpm
182-207 bpm

Therapies

All Rx Off
ATP During Charging, 35J x 6
All Rx Off
Burst(3), Burst(3), Burst(3), 35J x 3

Enhancements On: AF/Afl, Sinus Tach, Wavelet, Onset(Monitor), TWave, Noise(Timeout)

Clinical Status

Since 27-Jun-2025

Treated

VF	0
FVT (Off)	
VT	0
AT/AF(Monitor)	

Monitored

VT (Off)	
VT-NS (>4 beats, >182 bpm)	0
High Rate-NS	0
SVT: VT/VF Rx Withheld	0
V. Oversensing-TWave Rx Withheld	0
V. Oversensing-Noise Rx Withheld	0
AT/AF	266
Time in AT/AF	10.9 hr/day (45.3%)
Longest AT/AF	4 minutes

Functional

Last Week

Patient Activity	2.8 hr/day
------------------	------------

Cardiac Compass Trends (May-2024 to Jul-2025)