

Date: 2026-04-23
Request ID: 100048443

Microwave				
Element(s)	La, Na, Zr			
SOP#				
Acid Cocktail	10 mL H2O, 10 mL H2SO4			
Rack	7			
Vessel	PTFE			
Stir	90%			
Time (min)	Power (W)	T1 (C)	T2 (C)	P (bar)
15	1200	180	65	110
15	1500	240	65	100
30	1500	240	65	100
sample(s)	weight (g)	volume (mL)		
200127623_1	0.2041	200		
200127623_2	0.2052	200		

Date: 2026-04-23

Request ID: 100048443

sample	Dilution	conc. [mg/L]	La_ppm	%La
200127623_1	25/5	2.088	10230.28	1.02
200127623_2	25/5	2.107	10268.03	1.03

reported result: 10249.16 ppm

reported result: 1.02 %

oxide factor: 1.173

reported result: 1.20 %

lot: Manufacturer: Assurance, 28-70LAY exp: 2027-02-28

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200127623_1	1/1	5.692	5577.66	0.56
200127623_2	1/1	5.772	5625.73	0.56

reported result: 5601.69 ppm

reported result: 0.56 %

oxide factor: 1.348

reported result: 0.76 %

lot: Manufacturer: Assurance, 27-159NAY exp: 2026-08-30

sample	Dilution	conc. [mg/L]	Zr_ppm	%Zr
200127623_1	25/5	1.242	6085.25	0.61
200127623_2	25/5	1.309	6379.14	0.64

reported result: 6232.20 ppm

reported result: 0.62 %

oxide factor: 1.351

reported result: 0.84 %

lot: Manufacturer: Assurance, 28-64ZRY exp: 2026-10-30

Note(s): tested as received.