

Date: 2026-06-16
Request ID: 100048656

Microwave				
Element(s)	Ca			
SOP#				
Acid Cocktail	6 mL HCl, 2 mL HNO3			
Rack	7			
Vessel	PTFE			
Stir	90%			
Time (min)	Power (W)	T1 (°C)	T2 (°C)	P (bar)
15	1200	180	65	110
15	1500	240	65	110
30	1500	240	65	110
sample(s)	weight (g)	volume (mL)		
200128304_1	0.1053	250		
200128304_2	0.1078	250		

Date: 2026-06-16
Request ID: 100048656

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128304_1	100/5	1.951	92640.08	9.26
200128304_2	100/5	1.960	90909.09	9.09

ca result: 91774.58 ppm

ca result: 9.18 %

ca oxide factor: 1.399

ca oxide result: 12.84 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$