

Date: 2026-06-23

Request ID: 100048676

Hotplate		
Element(s)	Al, Co, Cu, Ti	
SOP#	FT wax digestion	
Cocktail	2 mL H2O, 10 mL 1:1 HCl, 10 mL 1:1 HNO3	
sample(s)	weight (g)	volume (mL)
200128352_1	0.2064	25
200128352_2	0.2104	25
200128353_1	0.2123	25
200128353_2	0.2179	25

Date: 2026-06-23

Request ID: 100048676

Al ICP analysis

sample	Dilution	conc. [mg/L]	Al_ppm	%Al
200128352_1	25/25	0.855	103.56	0.01
200128352_2	25/25	0.855	101.59	0.01

al result: 102.58 ppm

al result: 0.01 %

al oxide factor:

al oxide result: 0.00 %

al lot: *Manufacturer: Assurance, 27-171ALT exp: 2026-07-30*

Co ICP analysis

sample	Dilution	conc. [mg/L]	Co_ppm	%Co
200128352_1	25/25	0.045	5.45	0.00
200128352_2	25/25	0.045	5.35	0.00

co result: 5.40 ppm

co result: 0.00 %

co lot: *Manufacturer: Assurance, 28-73COY exp: 2026-10-30*

Cu ICP analysis

sample	Dilution	conc. [mg/L]	Cu_ppm	%Cu
200128352_1	25/25	0.000	0.00	0.00
200128352_2	25/25	0.000	0.00	0.00

cu result: 0.00 ppm

cu result: 0.00 %

cu lot: *Manufacturer: Assurance, 28-83CUT exp: 2026-09-30*

Ti ICP analysis

sample	Dilution	conc. [mg/L]	Ti_ppm	%Ti
200128352_1	25/25	0.030	3.63	0.00
200128352_2	25/25	0.030	3.56	0.00

ti result: 3.60 ppm

ti result: 0.00 %

ti lot: *Manufacturer: Assurance, 27-189TIT exp: 2026-08-30*

Note(s): *Duplicate results were inconsistent.*

Date: 2026-06-23

Request ID: 100048676

Al ICP analysis

sample	Dilution	conc. [mg/L]	Al_ppm	%Al
200128353_1	25/25	0.762	89.73	0.01
200128353_2	25/25	0.762	87.43	0.01

al result: 88.58 ppm

al result: 0.01 %

al oxide factor:

al oxide result: 0.00 %

al lot: *Manufacturer: Assurance, 27-171ALT exp: 2026-07-30*

Co ICP analysis

sample	Dilution	conc. [mg/L]	Co_ppm	%Co
200128353_1	25/25	0.033	3.89	0.00
200128353_2	25/25	0.033	3.79	0.00

co result: 3.84 ppm

co result: 0.00 %

co lot: *Manufacturer: Assurance, 28-73COY exp: 2026-10-30*

Cu ICP analysis

sample	Dilution	conc. [mg/L]	Cu_ppm	%Cu
200128353_1	25/25	0.000	0.00	0.00
200128353_2	25/25	0.000	0.00	0.00

cu result: 0.00 ppm

cu result: 0.00 %

cu lot: *Manufacturer: Assurance, 28-83CUT exp: 2026-09-30*

Ti ICP analysis

sample	Dilution	conc. [mg/L]	Ti_ppm	%Ti
200128353_1	25/25	0.029	3.41	0.00
200128353_2	25/25	0.029	3.33	0.00

ti result: 3.37 ppm

ti result: 0.00 %

ti lot: *Manufacturer: Assurance, 27-189TIT exp: 2026-08-30*

Note(s): *Duplicate results were inconsistent.*

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$