

Date: 2026-07-10

Request ID: 100048751

Katanax		
Analyte(s)	Ca	
SOP#		
Acid Cocktail	30 mL 1:1 HCl	
sample(s)	weight (g)	volume (mL)
200128537_1	0.2058	250
200128537_2	0.2082	250
200128538_1	0.2059	250
200128538_2	0.2031	250

Date: 2026-07-10
Request ID: 100048751

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128537_1	100/2	1.504	91350.83	9.14
200128537_2	100/2	1.515	90958.21	9.10

ca result: 91154.52 ppm

ca result: 9.12 %

ca oxide factor: 1.399

ca oxide result: 12.75 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

Date: 2026-07-10
Request ID: 100048751

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128538_1	100/2	1.507	91488.59	9.15
200128538_2	100/2	1.521	93611.52	9.36

ca result: 92550.05 ppm

ca result: 9.26 %

ca oxide factor: 1.399

ca oxide result: 12.95 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$