

Date: 2026-07-29

Request ID: 100048782

Hotplate		
Analyte(s)	Na	
SOP#		
Cocktail	50 mL 1:1 HCl	
sample(s)	weight (g)	volume (mL)
200128620_1	0.4889	100
200128620_2	0.4215	100
200128621_1	0.5078	100
200128621_2	0.5015	100
200128622_1	0.5055	100
200128622_2	0.5024	100
200128623_1	0.5066	100
200128623_2	0.5036	100
200128624_1	0.5092	100
200128624_2	0.5016	100

Date: 2026-07-29

Request ID: 100048782

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128620_1	25/4	0.884	1130.09	0.11
200128620_2	25/4	0.793	1175.12	0.12

na result: 1152.60 ppm

na result: 0.12 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s):

Date: 2026-07-29
Request ID: 100048782

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128621_1	25/4	0.966	1188.95	0.12
200128621_2	25/4	0.977	1216.97	0.12

na result: 1202.96 ppm

na result: 0.12 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s):

Date: 2026-07-29
Request ID: 100048782

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128622_1	25/4	0.933	1152.94	0.12
200128622_2	25/4	0.933	1160.06	0.12

na result: 1156.50 ppm

na result: 0.12 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s):

Date: 2026-07-29

Request ID: 100048782

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128623_1	25/4	0.840	1035.70	0.10
200128623_2	25/4	0.845	1048.08	0.10

na result: 1041.89 ppm

na result: 0.10 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s):

Date: 2026-07-29
Request ID: 100048782

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128624_1	25/4	0.840	1030.42	0.10
200128624_2	25/4	0.807	1005.53	0.10

na result: 1017.97 ppm

na result: 0.10 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s):

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$