

Date: 2026-07-27

Request ID: 100048804

Microwave				
Analyte(s)	Ca			
SOP#				
Acid Cocktail	6 mL HCl, 2 mL HNO3			
Rack	7			
Vessel	Quartz			
Stir	90%			
Time (min)	Power (W)	T1 (°C)	T2 (°C)	P (bar)
15	1200	180	65	110
15	1500	240	65	110
30	1500	240	65	110
sample(s)	weight (g)	volume (mL)		
200128674_1	0.1081	250		
200128674_2	0.1046	250		
200128675_1	0.1097	250		
200128675_2	0.1044	250		
200128676_1	0.1088	250		
200128676_2	0.1036	250		

Date: 2026-07-27
Request ID: 100048804

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128674_1	100/5	1.877	86817.76	8.68
200128674_2	100/5	1.907	91156.79	9.12

ca result: 88987.27 ppm

ca result: 8.90 %

ca oxide factor: 1.399

ca oxide result: 12.45 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

Date: 2026-07-27
Request ID: 100048804

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128675_1	100/5	1.946	88696.44	8.87
200128675_2	100/5	1.854	88793.10	8.88

ca result: 88744.77 ppm

ca result: 8.87 %

ca oxide factor: 1.399

ca oxide result: 12.42 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

Date: 2026-07-27
Request ID: 100048804

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128676_1	100/5	1.930	88679.53	8.87
200128676_2	100/5	1.837	88642.21	8.86

ca result: 88660.87 ppm

ca result: 8.87 %

ca oxide factor: 1.399

ca oxide result: 12.40 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$