

Date: 2026-07-30
Request ID: 100048834

Hotplate		
Analyte(s)	Na	
SOP#		
Cocktail	60 mL 1:1 HCl	
sample(s)	weight (g)	volume (mL)
200128746_1	0.1035	250
200128746_2	0.1081	250

Date: 2026-07-30
Request ID: 100048834

Na ICP analysis

sample	Dilution	conc. [mg/L]	Na_ppm	%Na
200128746_1	25/10	0.687	4148.55	0.41
200128746_2	25/10	0.712	4116.56	0.41

na result: 4132.55 ppm

na result: 0.41 %

na lot: *Manufacturer: Assurance, 27-159NAY exp: 2026-08-30*

Note(s): *Customer CofA shows 0.46% Na*

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$