

Date: 2026-08-07
Request ID: 100048869

Microwave				
Analyte(s)	Ca			
SOP#				
Acid Cocktail	6 mL HCl, 2 mL HNO3			
Rack	7			
Vessel	Quartz			
Stir	90%			
Time (min)	Power (W)	T1 (°C)	T2 (°C)	P (bar)
15	1200	180	65	110
15	1500	240	65	110
30	1500	240	65	110
sample(s)	weight (g)	volume (mL)		
200128818_1	0.1044	250		
200128818_2	0.1071	250		

Date: 2026-08-07
Request ID: 100048869

Ca ICP analysis

sample	Dilution	conc. [mg/L]	Ca_ppm	%Ca
200128818_1	25/1	2.133	127693.97	12.77
200128818_2	25/1	2.157	125875.35	12.59

ca result: #####

ca result: 12.68 %

ca oxide factor: 1.399

ca oxide result: 17.74 %

ca lot: *Manufacturer: Assurance, 28-35CAY exp: 2026-10-30*

Note(s):

$$\text{ppm } M+ = [\text{conc.}][\text{volume}][\text{dilution}]/[\text{weight}]$$

$$\%M+ = [\text{conc.}][\text{volume}][\text{dilution}]/([\text{weight}]*10,000)$$

$$\%MO = [\text{oxide factor}]*\%M+$$