

WET Labs, Inc. Philomath, OR 97370 (541) 929-5650

## ACS CALIBRATION RESULTS

Serial #:	acs135
Cal #:	001
Date:	4/23/13

Wavelength Range: 400 to 730 nm
---------------------------------

Tcal:	21.2	degrees Celsius
-------	------	-----------------

Ical:	21.0	degrees Celsius
-------	------	-----------------

Acal:	23.4	degrees Celsius
-------	------	-----------------

Temperature Range:	3.0	to	35.4	degrees Celsius
--------------------	-----	----	------	-----------------

Zero:	25.5	degrees Celsius
-------	------	-----------------

---

### Wavelength

Range: Wavelength range for which instrument use is recommended.

Tcal: Water temperature in flow tubes at time of water calibration.

Ical: Internal instrument temperature at time of water calibration.

Acal: Internal instrument temperature when recording air offsets in the \*.cal file.

Temperature Range: Internal instrument temperature range, for which instrument use is recommended.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.