

MERIDIAN GENOMICS CORE

High-Throughput DNA/RNA Sample Processing Manifest

Advanced Genomics Facility · Building B, Suite 400 · Apex Research Park

Project ID: PRJ-2026-NEX-0842

Intake Date: 2026-07-24

Batch Status: **QC PASSED - IN PREP**

Seq Platform: NovaSeq X Plus

Principal Investigator: Dr. Eleanor Vance (Apex Bio)
Primary Workflow: Whole Genome Sequencing (WGS)
Lab Technician: Marcus Thorne, CLS
Sample Count: 24 Samples (1 Plate, Wells A01-C08)

Submitting Org: Synapse Therapeutics Inc.
Target Reads/Cell: 30x Human WGS / 50k scRNA
Scheduled Run: 2026-07-26 (Lane 3 & 4)
QC Protocol: ISO/IEC 17025 Standard

1 Section 1: Sample Intake Checklist & Threshold Criteria

Technical Acceptance Criteria

Parameter	WGS Standard	scRNA-Seq
Min Concentration	$\geq 10.0 \text{ ng}/\mu\text{L}$	$\geq 500 \text{ cells}/\mu\text{L}$
Total Input Mass	$\geq 100 \text{ ng}$	$\geq 50,000 \text{ cells}$
Min Volume	$\geq 20 \mu\text{L}$	$\geq 15 \mu\text{L}$
A260/A280 Ratio	1.80 – 2.00	N/A
A260/A230 Ratio	2.00 – 2.20	N/A
RIN / DIN Score	DIN ≥ 7.5	RIN ≥ 8.0
Buffer Media	10mM Tris-HCl	1x PBS + 0.04% BSA

Intake Quality Assurance Log

- ✓ Sample tubes inspected for seal integrity and leakage.
- ✓ Barcodes scanned and matched against submission manifest.
- ✓ Fluorometric quantification (Qubit 4.0 BR DNA Assay).
- ✓ Microfluidic electrophoresis (Agilent TapeStation 4200).
- ✓ Cell viability assessed via automated trypan blue counting ($> 92\%$).
- ✓ Samples aliquoted into 96-well reaction plate (Lot #20260710).

2 Section 2: 96-Well Plate Batch Map (Plate Barcode: PLT-96-NEX-4412)

