

Certificate No: C358173

Group - Repeatability

Nominal Value: 20.0000048g

Group - Repeatability with a maximum value of 20.0000048g is not within 29.0% to 120% of max linearity 52.000007g

Certificate No: C359033

Weight - Lineality Up

Nominal Value: 75000.0g

Measured uncertainty 12.0g is not within required uncertainty 14.0g

Certificate No: C358451

No standard deviation present

Certificate No: C359245

Directions not present: North, South, East, West, Center, Down

No standard deviation present

Certificate No: C359246

Directions not present: North, South, East, West, Center, Down

No standard deviation present

Certificate No: C358853

Directions not present: West

Certificate No: C359251

Directions not present: North, South, East, West, Center, Down

No standard deviation present

Certificate No: C359117

Directions not present: North, South, East, West, Center, Down

No standard deviation present

Certificate No: C359255

Directions not present: North, South, East, West, Center, Up, Down

No standard deviation present

List of Certificates That Passed:

- C359249
- C353803
- C359107
- C359241
- C359048
- C358255
- C358854
- C354014
- C359106
- C355400

- C355707
- C358039
- C358855
- C358346
- C359041
- C359054
- C354788
- C353799
- C358173
- C358350
- C359122
- C353798
- C353802
- C353800
- C353804
- C353793