





\$809 563-0457 · 786 866-5606
\$ sales@phoenixcalibrationdr.com
\$ www.phoenixcalibrationdr.com
\$ Calle D, Esquina C. El Diamante
No. 18, Zona Ind. de Herrera
Santo Domingo, Rep.Dom. 11001
☐ □ □ ② □ @phoenixcalibrationdr



CERT. NO.: C343964

# CERTIFICATE OF CALIBRATION

### **GENERAL INFORMATION**

**Customer:** Cartones del Caribe, S.A.

**Contact:** Eduardo Feliz

**Service No.:** 51311 **Asset No.:** 548355

**Description:** Precision Balance

Manufacturer: Radwag
Model No.: WTC 600
Serial No.: 548355
Range: 0.5g to 600 g

**Location:** N/D

#### **CALIBRATION INFORMATION**

Calibration Date:09/06/2023Next Due:09/06/2024Cal. Freq.:12 MonthsAsFound Condition:In toleranceCalibration Results:Passed

Calibration Location: On-Site Calibration
Calibrated by: Pablo Romero
Environment: 71°F 54%RH
Cal. Procedure: DR-WI-0126

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NRC, NPL, etc.). This calibration conforms to the requirements of ISO/IEC 17025:2017. Results contained in this document relate only to the items tested or calibrated. The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, such that the confidence level approximates 95%. The customer is responsible for assigning the calibration frequency if needed, and it does not imply that the conformity of the equipment with the specifications is maintained. This certificate shall not be reproduced except in full and with the written consent of Phoenix Calibration. The statement of conformity is based on simple acceptance, whether the calibration result is within or outside the manufacturer's specification/acceptance limits, with the TUR expressed on the datasheet. It is the responsibility of the end user to determine if it is appropriate for your specific application. Pass: Results within limits. Fail: Results exceed limits.

#### **REMARKS**

Performed the routine of calibration, all evaluated parameters were found in Tolerance. The result is Passed.

## REFERENCE STANDARDS USED TO CALIBRATE EQUIPMENT

ID InstDescriptionSerial No.Cal. DateDue Date9959Weight Set 1mg to 10kgNSN08/19/202311/30/2023

## **REVISION COMMENT**

**SIGNATURES** 

Calibrated by: Pablo Romero

Quality Approval: Mayelin Cabrera

M. Cabrera

Date: 09/06/2023 Date: 09/12/2023







\$809 563-0457 · 786 866-5606 sales@phoenixcalibrationdr.com www.phoenixcalibrationdr.com Calle D, Esquina C. El Diamante No. 18, Zona Ind. de Herrera Santo Domingo, Rep.Dom. 11001 ff in ⊚ ☑ @phoenixcalibrationdr



CERT. NO.: **C343964** 

# **DATASHEET**

			Weig	ht - Desv Std				
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	0.0000	N/A	N/A	0.0000	N/A	**	n/a
			Weight	t - Excentricit	ty			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	0.0000	N/A	N/A	0.0000	N/A	**	n/a
			Weight - E	xcentricity C	enter			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
		,	Weight - Exc	centricity No	rtheast			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
	,			,				
		•	Weight - Exc	entricity Nor	thwest			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
		•	Weight - Exc	entricity Sou	ıtheast			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
		_						
				entricity Sou				
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound	NI/A	Meas. Uncert. **	TUR
g	N/A	200.00000	N/A	N/A	200.00	N/A	**	n/a
			Weight -	Linearity Do	wn			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	0.02	450.00000	449.98	450.02	450.00	<b>/</b>	0.0083	>10:1
g	0.02	300.00000	299.98	300.02	300.00		0.0083	>10:1
g	0.02	150.00000	149.98	150.02	150.00		0.0083	>10:1
g	0.02	0.00000	-0.02	0.02	0.00	N/A	**	n/a
								,
			Weight	: - Linearity U	Jp			
Units	Max. Error(Tol.)	Nominal	Low Limit	HighLimit	AsFound		Meas. Uncert.	TUR
g	0.02	0.00000	-0.02	0.02	0.00	N/A	**	n/a
g	0.02	150.00000	149.98	150.02	150.00	<b>/</b>	0.0083	>10:1
g	0.02	300.0000	299.98	300.02	300.00	<b>/</b>	0.0083	>10:1

End of datasheet

449.98

599.98

450.02

600.02

450.00

600.00

g

g

0.02

0.02

450.00000

600.00000

0.0083

0.0083

>10:1

>10:1







\$809 563-0457 · 786 866-5606
 \$sales@phoenixcalibrationdr.com
 \$www.phoenixcalibrationdr.com
 Calle D, Esquina C. El Diamante
 No. 18, Zona Ind. de Herrera
 Santo Domingo, Rep.Dom. 11001
 ™ © № @phoenixcalibrationdr



CERT. NO.: C343964

# **DATASHEET**

(✓) Compliance/Pass.

(X) Non-compliance/Fail.

(\*\*) Represents Non ISO/IEC 17025 Accredited Values.