

Bid Document

Bid Details	
Bid End Date/Time	23-07-2020 12:00:00
Bid Opening Date/Time	23-07-2020 12:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Human Resource Development
Department Name	Department Of Higher Education
Organisation Name	Mizoram University
Office Name	Aizawl Mizoram
Total Quantity	1
Item Category	kjeldahl digestion cum distillation assembly unit with fume hood
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Kjeldahl Digestion Cum Distillation Assembly Unit With Fume Hood (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Detailed specification	Digestion and distillation combined unit	yes	*
	Fume Hood	Available	*
	Heater	Mantle Type (2 sets, One for Digestion and one for distillation)	*
	Unit supplied	With glass parts	With glass parts, With out glass parts
	Construction	MS Powder coated	*
	Heating capacity of the heater (Watt)	500	*
	Units of test	6	*
	Each heating unit fitted with individual energy regulator	yes	*
	Supplied complete with condenser rack,connecting tubes and clamps	yes	*
	Power	230V plus minus 10 percent, 50 Hz	*
	Flask capacities (ml)	300	*
	Weight of the complete unit (grams)	5.4	*
	Details of other accessories to be supplied along with the unit	-	*
Test report Details	Availability of Type Test Report from Central Govt. /NABL/ILAC Accredited lab. To prove conformity to the specifications.	yes	*
	Type Tests report to be furnished to the buyer on demand	yes	*

Name of Lab	-	*
Type Test report No.	-	*
Address of Lab	-	*
TTC report date	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	F.KAPMAWII	796004,Mizoram University, Tanhril, Aizawl, Mizoram	1	30

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
2. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
3. ISO 9001: The bidder must have ISO 9001 certification.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---