

RA Document

RA Details	
RA Technical End Date/Time	20-07-2020 11:00:00
RA Opening Date/Time	20-07-2020 11:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Gujarat
Department Name	Health & Family Welfare Department Gujarat
Organisation Name	N/a
Office Name	Health Branch District Panchayat Junagadh
Total Quantity	100
Item Category	bacterial or viral filters without heat and moisture exchanger/hme
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

RA splitting not applied.

Bacterial Or Viral Filters Without Heat And Moisture Exchanger/hme (100 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Packing	Type of Packing	Blister peelable pouch	*
Configuration of Product	Housing material of the Filter	Medical grade Poly propylene	Medical grade Poly propylene, Medical grade PVC
Features	Viral Filtration Efficiency (%)	99.99	*
	Appropriate for Tidal Volume range (ml)	150 ml to 1500 ml	150 ml to 1000 ml, 150 ml to 1500 ml
	Total Shelf life (in years)	3	1 - 5
	Mode of filtration	Mechanical,Electrostatic Media	*
	Fitting design	Straight	Straight, Angled
	CO2 Port	No	Yes, No
	Size of bacteril/viral filter	Adult	Adult, Mini/Paediatric
	Latex Free	Yes	*
	Sterlized	Yes	*
	Filtration Method	Electrostatic media	Electrostatic media, Hydrophylic media
	Fitting Size(Inlet and outlet size) (as per ISO)	22 IDx15ID/22OD	22 IDx15ID/22OD, 15F/22M - 15M/22F, 10mmM - 10mmM
	Dead space	35	*
	Disposable	Yes	*
Certifications	WHO-GMP or GMP certified	No	*
	CE marked	No	*
	Firm Certified to ISO 9001:2008 or latest	Yes	Yes, No
	Other Certifications	-	*
	ISI Marking	No	*
	CML No	-	*
Flow rate wise resistance	Resistance to flow @ 30L/min (in mmH2O)	6	*
	Resistance to flow @ 60L/min (mmH2O)	0.1	*
	Resistance to flow @ 90L/min (mmH2O)	1	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	Radadiya Harsukhbhai Fulabhai	362001,HEALTH BRANCH DISTRICT PANCHAYAT OPPOSITE SHASHI KUNJ NEAR SARDAR BAUG	100	30	N/A

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---