

Bid Document

Bid Details	
Bid End Date/Time	23-07-2020 11:00:00
Bid Opening Date/Time	23-07-2020 11:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Human Resource Development
Department Name	Department Of Higher Education
Organisation Name	Mizoram University
Office Name	Aizawl Mizoram
Total Quantity	2
Item Category	fume hoods for laboratory with cupboard
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Fume Hoods For Laboratory With Cubboard (2 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
DIMENSIONS, FEATURES	Size of Sink (mm)	0	*
	Power Rating	-	*
	Drawing No.	-	*
	Product Code available on Web	-	*
	Safety Device Trip (Nos)	-	*
	Material of Base Cabinet	-	*
	Working Platform Material	-	*
	Double wall Construction	Yes	*
	Ducting Material	-	*
	Details of Electrical outlets	-	*
	Product Code No	-	*
	Air Flow (Cubic metre/hr)	100	*
	Width of Base Cabinet (mm)	900	10 - 10000
	Material of Front shutter	Toughned Float Glass	Toughned Float Glass, other, Epoxy Coated Steel
	Overall lenth of Fumehood (mm)	1200	100 - 5000
	Base Cabinets	Without	With, Without
	Ducting for hood	Without	With, Without
	Efficiency & Noise level of Blower (%, dB)	-	*
	Thickness of wall (mm)	10	0.91 - 500
	Operating Temperature of Fume hood (degree C)	40	*
	Material of Fumehood	-	*

No. of Drawers in the Base Storage Unit (Nos)	0	*
Fumehood installation Kit and Accessories	With	With, Without
Height of Base Cabinet (mm)	750	1.0e-06 - 9999999
Type of Filter	-	*
Type of Fumehood	Laboratory	*
Length of Base Cabinet (mm)	1200	10 - 5000
Fire Resistant	Yes	Yes, No
Utility Pipe lines for Nitrogen,Compressed Air & Water etc.	No	Yes, No
Face Velocity (m/s)	0.45	*
Type of Fumes suitable for	-	*
Overall Width of Fumehood (mm)	900	*
Design of base Cabinet	-	*
Wattage of Blower (W)	-	*
Material of sink	-	*
Overall Height of Fumehood (mm)	2300	*
Type of Blower	-	*
Design details of baffles	-	*
Design details of Electrical Services	-	*
No. of Electrical Socket and Switch (Nos)	-	*
Material of the Blower	-	*
Working Platform Thickness (mm)	-	*
Provision of Electrical Outlets	Yes	Yes, No
Drawing available on Web	Yes	Yes, No
Features of various Utilities	-	*
Fumehood Baffles	Without	With, Without

Front shutter Fumehood	With	With, Without
Material of Insulation	-	*
Type of Base cabinets	Movable with Wheels	Movable with Wheels, Movable without Wheels
Non Corrosive treatment of inside compartment	Yes	*
Availability of Sink	Without	With, Without
No. of Shutters (Nos)	1	1, 2, 3, 4, 5, 6, 7, 8, 9, 10
No. of Shelves in Base storage units (Nos)	0	*
Materials of Baffles	-	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	F.KAPMAWII	796004,Mizoram University, Tanhril, Aizawl, Mizoram	2	30

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
2. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
3. ISO 9001: The bidder must have ISO 9001 certification.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---