

Bid Document

Bid Details	
Bid End Date/Time	22-07-2020 15:00:00
Bid Opening Date/Time	22-07-2020 15:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	15 (Days)
Ministry/State Name	Jammu & Kashmir
Department Name	Health And Family Welfare Department Jammu And Kashmir
Organisation Name	Government Medical College
Office Name	Doda
Total Quantity	1
Item Category	Hot air oven (PAC Only)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Hot Air Oven (1 pieces) (Under PAC)

Make	sciencetech
Model	se-127

Technical Specifications

* [As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE	*
	Certification number and date	.2021 CE-7190 DATE: 26.03.2018 EXPIRY DATE: 25.03	*
	Confirming Standards for Electrical Safety	IEC- 60601 with latest amendments or IS 13450 with latest amendments	*
	Compliance of Quality Systems for medical devices	ISO 13485	*
Performance Parameters	Purpose	Hot Air Oven is commonly used for dry heat sterilization which is a method of controlling micro-organisms	*
	Capacity in litres	101-150	101-150
	Power ratings in watts	2000	2000
	Should provide 3 heating elements on 3 sides of equipment to ensure uniform temperature on all shelves	Yes	*
	Ambient temperature range in °C	50 °C to 200 °C	50 °C to 200 °C
	Temperature Accuracy in °C	±1°C	±1°C
	Insulation	Glass wool	Glass wool
	Controller	Microprocessor based PID control	*
	Type of Display	LCD	LCD
	Parameters displayed	Time, Temperature settings and temperature reading	*
	Type of Temperature sensor	PT 100	*

	Power supply	220 V , 50 Hz Single Phase	*
	Door shall be insulated	Yes	*
	Door shall be fitted with heavy hinges	Yes	*
	Availability of Double glass window on door	No	Yes, No
	Door should have locking facility	Yes	*
	Type of Shelves	Removable	Removable
	Number of shelves	3	3
	Air Circulation	Motorized Blower for forced air circulation	*
	Blower Motor power in HP	0.5	*
	Blower motor shall be totally enclosed	Yes	*
	Blower motor shall be air cooled	Yes	*
	Weight in kgs	55	*
	Length of power cord fitted with compatible Indian plug in meters	3	*
	Operating temperature and humidity	10 to 40 degree centigrade and 15-90% humidity	*
	Over temperature protection	Yes	*
	Safety alarm system	Yes	*
	Instruction manual to be provided	Yes	*
	Oven should be corrosion resistant	Yes	*
	Warranty in Years	2	2
Material Parameters	Material of Inner Chambers	SS 304	*
	Material of Outer Chamber	MS with powder coated paint	MS with powder coated paint
	Material of Shelves	SS wire mesh	SS wire mesh
Reports	ISO certification for manufacturer	ISO 9001	*
	ISO certification date and number	7186 DATE: 26.03.2018 EXPIRY DATE: 25.03.2021	*

	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*
--	--	-----	---

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Vijay Katoch	182202,GMC DODA NEAR D.C OFFICE	1	30

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods
2. ISO 9001: The bidder must have ISO 9001 certification.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---