

### Bid Document

Bid Details	
Bid End Date/Time	22-07-2020 15:00:00
Bid Opening Date/Time	22-07-2020 15:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Ministry Of Power
Department Name	Contracts And Materials
Organisation Name	Ntpc Ltd
Office Name	Jharkhand
Total Quantity	5
Item Category	glucose monitors or meters-Glucometer
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No
Estimated Bid Value	11250

#### EMD Detail

Required	No
----------	----

#### ePBG Detail

Required	No
----------	----

#### Splitting

Bid splitting not applied.

### Glucose Monitors Or Meters-Glucometer ( 5 pieces )

#### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
COMPLETE SET	Availability of Glucostrip & Lancet in open market all over India	Yes	*
	Operating Relative humidity (%)	15 to 90	*
	<b>Number of sets of Glucostrip and lancet in a bundle</b>	25	25, 50, 100, 200 Or higher
	<b>Availability of safety certifications - FDA (US)/CE (EU) -- BIS/ISO:13485 &amp; ISO: 15197</b>	CE (EU)	FDA (US), CE (EU)
	<b>Size of LCD display digits (mm)</b>	70	70.0 - 100.0 Or higher
	Inbuilt software for accurate measurement	Yes	*
	Auto-shut off of device when not in use	Yes	*
	Needle fully shielded before and after use	Yes	*
	<b>Accuracy of device (%)</b>	5	0.0 - 10.0 Or lower
	Calibre stick appropriate for 500 tests	Yes	*
	Number of tests on one battery (No. of Test results)	1000	*
	<b>Weight of device</b>	50	10.0 - 50.0 Or lower
	Operating Temp	0 to 55 Degree Celcius	*
	<b>Warranty of device (Years)</b>	3	3
	<b>Reproducibility of device (%)</b>	5	0.0 - 5.0 Or lower
	Facility for Data transfer through USB device	Yes	*
	Operating Manual of device	Yes	*

Single use pressure activated autodisable Safety Lancet	Yes	*
Technology of device	Electrochemical	*
<b>Size of blood sample (Micro litre)</b>	0.5	0.0 - 1.0 Or lower
<b>Shelf life of gluco strips (Months) at the time of supply</b>	12	More than 12.0 Or higher
<b>Penetration depth to suit capillary requirement</b>	Adjustable	Adjustable
Retraction of needle into casing after sampling	Automatic	*
Protection from blood borne infections	Yes	*
Number of cells (nos)	1	*
Display of time & date	Yes	*
<b>Reading time of device</b>	5	0.001 - 5.0 Or lower
Display of sugar in mg/dl	Yes	*
Automatic code detection of Glucostrip	Yes	*
<b>Life of Gluco-strips after opening of container</b>	3	3.0 - 60.0 Or higher
Dimensions of the Device (Length(L) xWidth(W) x Height (H))	-	*
Range/Linearity of device (mg/dl)	30 to 700(Min.)	*
<b>Memory capacity of device (Test results)</b>	500	500.0 - 9999999.0 Or higher
Whole device should be cleanable and sterilizable with alcohol	Yes	*
Control solution/Calibre for 500 tests	Yes	*
Availability Of Type Test Reports(TTRs) of the Product From Central Govt/NABL/ILAC Accredited Lab Showing Compliance To The Specification, If	Yes	*

	TTRs available must furnish details (No, Date and Name of Laboratory) of the same		
	Date of test report by laboratory	-	*
	Test report number	-	*
	Name, address and contact number/email of the laboratory	-	*

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

### Additional Specification Documents

### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Renu Kumari	829119,ADM Building, PVUNL, P.O.- Patratu, Distt- Ramgarh	5	30

### Bid Specific Additional Terms and Conditions

1. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
2. Scope of supply (Bid price to include all cost components) : Only supply of Goods
3. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
4. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
5. Warranty period of the supplied products shall be 3 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---