

Bid Document

| Bid Details | |
|--|---|
| Bid End Date/Time | 16-07-2020 12:00:00 |
| Bid Opening Date/Time | 16-07-2020 12:30:00 |
| Bid Life Cycle (From Publish Date) | 90 (Days) |
| Bid Offer Validity (From End Date) | 60 (Days) |
| Ministry/State Name | Ministry Of Power |
| Department Name | Na |
| Organisation Name | N/a |
| Office Name | Nhpc Limited, Tanakpur Power Station, Banbassa, 262310, Uttarakhand |
| Total Quantity | 1 |
| Item Category | Automatic External Defibrillator (AED) |
| MSE Exemption for Years of Experience and Turnover | No |
| Startup Exemption for Years of Experience and Turnover | No |
| Bid to RA enabled | No |
| Inspection Required | No |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

Automatic External Defibrillator (AED) (1 pieces)

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|-----------------------|---|--|--|
| Product Configuration | Clinical Purpose | Automatic external defibrillator should be fully automated, using biphasic technology as per manufacturer recommendations energy level (joules) to deliver shock energy, complying with latest guidelines of AHA and must be upgraded to any future changes. | * |
| | Display | AED have an LCD screen that display text prompts & ECG on screen | AED have an LCD screen that display text prompts & ECG on screen |
| | Alarms/Prompts: | Must provide both visual and audible prompts on CPR complying with latest AHA guidelines. | * |
| | Construction: | Compact, light weight, portable with easily identifiable ON/OFF switch and preferably with a handle to carry | * |
| | IPX rating for dust & water protection: | IP55 or better | * |
| Power | Power Source: | AED shall be battery operated | * |
| | Battery: | AED Battery shall be non rechargeable, disposable and easily replaceable battery. | * |
| | Battery Life: | Standby life of AED battery shall be of \geq 4 years, sufficient for eight hours of monitoring or minimum 100 shocks | * |
| Accessories | Disposable electrode pads: | Adult Disposable Electrode Pads - 2 nos. and Paediatric Disposable electrode pads - 2 Nos. are to be supplied for each unit of AED. If the Buyer demands, Seller shall | * |

| | | | |
|--------------------------------------|---|---|--|
| | | supply total 4 nos. of Disposable Electrode Pads suitable for Adult /Paediatric patients as per the requirement of the Buyer. | |
| Certification & Warranty | Warranty of AED: | 5 years, [With provision of free replacement of Battery during warranty period] | 5 years, [With out provision of free replacement of Battery during warranty period], 5 years, [With provision of free replacement of Battery during warranty period] Or higher |
| | Shelf life of Disposable electrode pad: | Two Years (if unused) | * |
| Operating Conditions & Certification | Operating Temperature: | 0 degree to 50 degree C | * |
| | Operating Humidity: | 12% to 90% RH, non condensing | * |
| | Operating Altitude: | 100 to 15000 ft or above | * |
| | Certification | - | * |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|--|----------|---------------|
| 1 | Dhirendra Pratap Patel | 262310,NHPC Limited. Tanakpur Power Station, Banbassa-262310 | 1 | 30 |

Special terms and conditions for category Automatic External Defibrillator (AED)

1. Comprehensive warranty Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares,. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables .Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.Service centresDetails of Service outlets in India to render services for equipment to be furnished to buyer/consignees

with complete address ,telephone numbers, e mails etc at time of making the supplies .It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled .Details of toll free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.Source of supplyIt shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.Packing and MarkingMedical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transshipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take in to consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed .Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date , brief description of goods including quantity ,. Packing list reference number , country of origin of goods and any other relevant details.Spare PartsSeller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM . It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies .In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available.OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.Installation, Training, ManualsSeller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and Supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations , training and manuals the same shall also be applicable. Electrical safety checkingSellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee. They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call. Software All software updates should be provided free of cost during warranty period.

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---