

Bid Document

Bid Details	
Bid End Date/Time	21-07-2020 15:00:00
Bid Opening Date/Time	21-07-2020 15:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Delhi
Department Name	Health And Family Welfare Department Delhi
Organisation Name	N/a
Office Name	Ambedkar Nagar Hospital Dakshin Puri Delhi 110062
Total Quantity	1
Item Category	Arterial Blood Gas Analyser
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Bidder Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Arterial Blood Gas Analyser (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
---------------	--------------------	--------	----------------------------------

Standards	Certification of the product	-	*
	Certification number and date	-	*
	Confirmity to Manufactures Certification	ISO 9001 & ISO 13485	*
Performance Parameters	Blood Gas Analyser (BGA) System	Blood gas analyser is an equipment for measuring and calculating complete Blood Gas Analysis	*
	Type of blood gas analyser	Hand held	Hand held
	Type of technology	Cartridge (single use)	Cartridge (single use)
	Parameters measured by blood gas analyser	pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , pH and calculated parameters like HCO ₃ ⁻ , Base excess, TCO ₂	*
	Additional Parameters measurable by the blood gas analyser	Cl-,iCa,Glu,BUN,Hct,Creatinine,Lacate	Cl-, iCa, Glu, BUN, Hct, Creatinine, Lacate
	Cardiac Parameters measurable	ACT, PT/INR, Troponin, BNP, CKMB	ACT, PT/INR, Troponin, BNP, CKMB
	Neonatal Parameters	Not available	nBil, Not available
	Monitor type	LCD	LCD
	Type of printer	External Thermal Printer	External Thermal Printer
	Cartridge / reagent to be used can be stored mostly at room temperature between 2-30 Degree C, Automatic calibration (hint: seller shall inform the exact details to buyer at the time of making the supplies)	Yes	Yes
	"Shall have patient Bar Code facility, Ability to run a wide range of patient panels on one system"	Yes	Yes
	Capability to transfer patient data to LIS/HIS	Yes	Yes
	Weight in kg	≤ 1 kg for hand operated	≤ 1 kg for hand operated
	Sample Volume in micro litre	100	*
	Power source	Rechargeable battery	Rechargeable battery
	Battery backup provided	Through rechargeable battery for 30 mins	Through rechargeable battery for 30 mins
	Shelf life of the reagent / cartridges as applicable in months	6	6
	On-board shelf life of reagent in weeks	24	2, 4, 6, 8, 24, Not applicable
	Warranty (including electrode and membrane etc as applicable for the analyser offered) in years	5	5

	User/Technical/Maintenance manuals to be supplied in original in English in hard and soft copy	Yes	*
	Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes	*
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes	*
Reports	Availability of ISO certification to manufacturer	ISO 13485	*
	ISO certification date and number	-	*
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Sudhir Kumar	110062, AMBEDKAR NAGAR HOSPITAL, DAKSHIN PURI, NEW DELHI	1	2

Bid Specific Additional Terms and Conditions

- Provisions for workmen compensation, insurance, public liability, property damage, insurance etc. to be ensured by the contractor. Also, safety measures prescribed for working in Extra High Voltage & use of personnel protective equipment must be adhered to.
- IMPORTED PRODUCTS:** In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
- Upload Manufacturer authorization:** Wherever Authorised Distributors are submitting the bid, Manufacturers Authorisation Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.
- The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
- Over and above the normal Warranty terms as per GeM GTC, the successful bidder / OEM shall have to provide Comprehensive Warranty during the entire Standard warranty period as per contract. : The comprehensive warranty shall be covering the following scope ALL PARTS (Upload an undertaking with the bid confirming compliance by the bidder if Bidder is taking onus of this compliance. In case OEM is taking onus of this compliance, OEM undertaking is to be uploaded along with Bidder undertaking)
- Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

8. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---