



Bid Document

Bid Details				
28-07-2020 13:00:00				
28-07-2020 13:30:00				
90 (Days)				
80 (Days)				
Ministry Of Health And Family Welfare				
Department Of Health And Family Welfare				
Post Graduate Institute Of Medical Education And Research Chandigarh				
Pgimer				
1				
Deep freezer				
No				

EMD Detail

Advisory Bank	State Bank of India
EMD Percentage(%)	0.50
EMD Amount	3000

ePBG Detail

11	
I I Doguirod	l No
Required	INO
11 '	

- (a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.
- (b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

Professor

Pgimer, Department of Health and Family Welfare, Post Graduate Institute of Medical Education and Research Chandigarh, Ministry of Health and Family Welfare (Bikash Medhi)

Splitting

Bid splitting not applied.

Deep Freezer (1 pieces)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number)	CE (with 4 digit notified number)
	Certification date and number	-	*
	Conformity to electrical safety	IEC 61010-1, IEC 61010-2-40 or IS 13450	*
Performance Parameters	Purpose	For storage of various biological products including ATCC cultures, enzymes, vaccines, chemicals or material testing components for a longer period of time	*
	Capacity in liters	501-600	501-600
	Type of cabinet	Vertical	Vertical
	Ground Clearance in mm	100 mm	*
	Cooling type	Direct cooling	*
	Freezer body	Galvanised steel body with tough powder coated exterior finish	*
	Castors	Heavy duty lockable	Heavy duty lockable
	Lockable outer doors and lids	Yes	*
	Material of internal cabinet	SS	*
	Freezer should use CFC- free, HCFC-free non flammable refrigerants	Yes	*
	Refrigeration system must be energy efficient	Yes	*
	Pull down time in hrs	5.2 hrs	*
	Operating Temperature in degree celcius	minus 10 degree to minus 40	*
	Increment variation setting	1° C	*
	Number of corrosion resistant racks	3	3
	Control	Fully programmable microprosseor controlled with membrane keypad and eye level control panel	*
	Number of inner stroage compartments with insulated doors and	4	4

	Digital temperature display	Yes	*
	Should have battery back up for the display and security lock for the display	Yes	*
	Alarms	Audible and visible alarm system for unwanted temperature rise, sudden power failure, system failure and battery low	*
	Noise level in dB	55	*
	Energy consumption rate (@KWh/day)	8.5	*
	Stabilizer should be capable to run any voltage between 190V - 270 V	Yes	*
	Warranty of Stabilizer in Years	5 years	*
	Warranty of Freezer in Years (from the date of installation)	3	3
Additional Parameters	Freezer must have washable condenser filter indicator which should keep fins free of dust to maintain peak cooling efficiency	Yes	*
	One 5KVA servo stabilizer to be supplied with freezer	Yes	*
	CO ₂ cylinder should be supplied with freezer for backup	No	No
	Tempearture Chart recorder to be provided	No	No
Reports	Manufacturer should have ISO certification	Yes	*
	ISO certification date and number	-	*
* Considirations bisblishbad i	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Bikash Medhi	160012,Department of Pharmacology, POST GRADUATE INSTITUTE OF MEDICAL EDUCATION AND RESEARCH, sector 12 CHANDIGARH	1	30

^{*} Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Bid Specific Additional Terms and Conditions

- 1.If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
 2.Scope of supply (Bid price to include all cost components) : Supply Installation Testing and Commissioning of Goods

This Bid is also governed by the General Terms and Conditions

---Thank You---