

Bid Document

Bid Details	
Bid End Date/Time	23-07-2020 13:00:00
Bid Opening Date/Time	23-07-2020 13:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	65 (Days)
Ministry/State Name	Uttar Pradesh
Department Name	Medical Education Department Uttar Pradesh
Organisation Name	Dr. Ram Manohar Lohia Institute Of Medical Sciences
Office Name	Lucknow
Total Quantity	150
Item Category	Patient Bed (Fowler and Semi Fowler)
Bidder Turnover (Last 3 Years)	14 Lakh (s)
OEM Average Turnover (Last 3 Years)	108 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years Of Experience and Turnover	Yes
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance,Bidder Turnover,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	10 %
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Advisory Bank	State Bank of India
EMD Percentage(%)	2.00
EMD Amount	54000

ePBG Detail

Advisory Bank	State Bank of India
ePBG Percentage(%)	5.00
Duration of ePBG required (Months).	15

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

Director

Dr. Ram Manohar Lohia Institute of Medical Sciences, Vibhuti Khand, Gomti Nagar, Lucknow-226010
(Prof. Nuzhat Husain)

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Patient Bed (Fowler And Semi Fowler) (150 pieces)**Technical Specifications**

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Performance Parameters	Type of bed	Fowler	Fowler
	Number of bed platform sections	4 for fowler bed	4 for fowler bed
	Availability of Backrest section	Yes	*
	Backrest section linkage	welded with MS CRCA tube (14 G)	*
	Availability of Leg section	Yes	Yes
	Leg section linkage	welded with MS CRCA tube (16 G)	welded with MS CRCA tube (16 G)
	Mattress platform shall be perforated	Yes	*
	Screw mechanism	welded with ERW MS tube (18 G) in M.S. cover made from (16 G) ERW tube	*
	Manual adjustments	Backrest , knee rest through two screw systems with thrust bearings individually manouvred by two handles for fowler bed	Backrest , knee rest through two screw systems with thrust bearings individually manouvred by two handles for fowler bed
	Back rest - min 45 % of the frame length	Yes	*
	The bottom end of the H legs are provided with PVC shoes	NA in case of bed with castors	Yes in case of bed without castors, NA in case of bed with castors
	The bed has head & foot panels detachable by hand without need of any tool	Yes	*
	Saline rod provision	four locations on the bed platform to hold stainless steel Telescopic Saline rod 12mm dia with 15.87 mm dia x 1.2 mm (18 G) stainless steel	*
	Provision for mounting of syringe pump	outer covering tube with a knob to mount syringe pump	*

Finishing & workmanship shall be of high standard	Yes	*
All corners shall be rounded off so that there shall be no sharp corners and holes should be burr free	Yes	*
Coating for metal parts	M.S. tubular parts, linkages, flats are to be In-house, pretreated / shot blasted and Epoxy powder coated with coating thickness 50 to 100 microns	*
Material of handle	SS	SS, Plastic, MS
Number of detachable moulded PP with PVC insert head and foot side panels	4	*
Provision of castors	H' Type Legs with 600 mm Height with 125 mm(5") dia High Grade Synthetic Body Castors, two with brake, two w/o brake	H' Type Legs with 600 mm Height with 125 mm(5") dia High Grade Synthetic Body Castors, two with brake, two w/o brake
Number of hooks in Telescopic I V Pole	4	2, 4
Urine Bag Holder	Yes	*
Moulded Chart Holder	Yes	*
Provision of mattress	Yes	*
Number of sections in mattress	4 for fowler bed	4 for fowler bed
Provision of collapsible Railing	Yes	Yes
Material and Dimensional Parameters	Overall Length of bed in mm with 5% tolerance	2050 to 2100
	Overall width of bed in mm with 5% tolerance	900 to 950
	Size of Buffer in mm	100
	Bed Platform frame size with 5% tolerance	2070 mm L x 910 mm W
	Mattress Platform size with 5% tolerance	1970 mm L x 790 mm W
	Buffer to Handle Length with 5% tolerance	2275 mm

Fixed Height	600 mm	*
Material and size of section	1.2 mm (18 G) CRCA M.S perforated sheet top	*
Overall length of Back Rest	775-800	775-800
Overall width of back rest	750-775	*
Overall length of Knee Rest	350-400	350-400
Overall width of Knee Rest	750-775	750-775
Overall length of Leg Rest	550-600	*
Overall width of Leg Rest	750-775	750-775
Overall length of Fix Section	100-150	100-150
Overall width of fix section	750-775	750-775
Raised Backrest Angle	65°	*
Raised Kneerest Angle	40°	*
Bed frame tube size	60 mm x 30 mm x 1.6 mm (16 G) Thick ERW tube	*
Dimensions for Fitting of frame	ERW tubes diameter 31.75 mm x 1.2 mm (18 G) thick and diameter 22.22 mm x 1.2 mm (18 G) thick and MS sheet having thickness 3.0 mm (11 G)	*
Diameter of Four corner rubber buffers in mm	75	*
Material of head and foot panels	ABS Plastic	*
Diameter of MS tubular in mm	30 to 50 mm	*
Diameter of SS Laminated	30 to 50 mm	*
Patient Working Load	135 Kg	*
Safe Working Load	150 Kg	*
Thickness of PU foam of mattress in inch	4	*

	Density of PU foam of mattress in Kg/m3	32	*
	Material of mattress cover of mattress	PVC	*
	Colour of mattress cover	Blue	Blue, Black, Maroon, NA (incase mattress is not provided)
	Material of collapsing railing	SS	SS
	Types of collapsing railing	Collapsible	Collapsible
Reports	Manufacturer Certificates	ISO 9001:2015	*
	Details of manufacturer certificates such as number and date	-	*
	Availability of certificates of quality system	ISO 13485	*
	Availability of test report from govt/ NABL/ILAC accredited lab covering all parameters	No	Yes, No
	Test report number , Incase test reports not available , put NA	-	*
	Test report date, Incase test reports not available , put NA	-	*
	Name of test lab , Incase test reports not available , put NA	-	*
	Address of test lab , Incase test reports not available , put NA	-	*
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	Yes, NA ,Incase test reports not available

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Mohit Kumar	226010,Vibhuti Khand, Gomti Nagar, Lucknow	150	60

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
2. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
3. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
4. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
5. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
6. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
7. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
8. ISO 9001: The bidder must have ISO 9001 certification.
9. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
10. Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid

11. For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:
- a. Purchase Order copy along with Invoice(s) with self-certification by the bidder that supplies against the invoices have been executed.
 - b. Execution certificate by client with order value.
 - c. Any other document in support of order execution like Third Party Inspection release note, etc.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---