

Bid Document

Bid Details	
Bid End Date/Time	24-07-2020 18:00:00
Bid Opening Date/Time	24-07-2020 18:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Ministry Of Steel
Department Name	Nmdc Limited
Organisation Name	Nmdc Limited
Office Name	Hyderabad
Total Quantity	1
Item Category	Density meter
Bidder Turnover (Last 3 Years)	1 Lakh (s)
Experience Criteria	3 Year (s)
MSE Exemption for Years Of Experience	Yes
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,Past Performance *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Past Performance	10 %
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Density Meter (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
TYPE/APPLICATION	Type	Bench-Top	Bench-Top
PERFORMANCE	Measuring technique	Oscillating U-Tube	*
	Measuring range of density(g/cm ³)	3	*
	Resoloution(g/cm³)	0.0001	1.0E-4
	Accuracy of density measurement(g/cm³)	0.01	0.0001, 0.001, 0.01 Or lower
	Repeatability Standard Deviation in density(g/cm ³)	0.0002	*
GENERIC	Minimum Sample Volume	5	*
	Display	In Built display	In Built display
	Touch display	No(NA for External Display)	No(NA for External Display)
	Display size	2	*
	Connectivity interface	RS 232	*
	Housing Material	Poly carbonate	*
	Net Weight	-	*
	Overall Dimension(W x D x H) (mmxmmxmm)	-	*
	URL of the uploaded	-	*

	product catalogue		
POWER SOURCE	Power Source	Mains (230 ± 10 Volts / 50 Hz AC)	*
	Battery Size, If Non Rechargeable	NA	*
	Battery Chemistry, if Rechargeable Battery	NA	*
	Number of Batteries required for operation	0	*
	Number of Batteries Inclusive in the Scope of Supply	0	*
	Battery capacity (mAh)	0	*
	Batteries required for operation are inclusive in the scope of supply	No	*
	Battery Life	0	*
FEATURES	Salient Features	Viscosity correction, bubble detection, Temperature correction	*
ACCESSORIES	Accessories in the scope of supply	Suitable AC Adaptor	*
WARRANTY	Warranty on Equipment	12	12, 24 Or higher
OPERATING ENVIRONMENT/ENCLOSURE PROTECTION	Minimum Operating Temperature	-1	*
	Maximum Operating Temperature	60	*
	Operating Humidity	80	*
	Enclosure Protection rating	IP 65	*
CALIBRATION CERTIFICATE	Calibration Certificate provided	Yes	*
	Calibration Certificate issuing authority	OEM Certificate	*
TEST REPORT	Type of lab which carried out Test of Complete Product to prove the conformity of product as per specification	OEM	*
	Test Report Number and Date	-	*
	Name and address of the Lab where Test conducted	-	*

- * Specifications highlighted in bold are the Golden Parameters.
- * Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Tadiparthi Jayakumar	500007,NMDC LTD R AND D UPPAL HYDERABAD	1	30

Bid Specific Additional Terms and Conditions

1. ISO 9001: The bidder must have ISO 9001 certification.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---