

### Bid Document

Bid Details	
Bid End Date/Time	21-07-2020 16:00:00
Bid Opening Date/Time	21-07-2020 16:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Ministry Of Labour And Employment
Department Name	Employees State Insurance Corporation
Organisation Name	Employees State Insurance Corporation
Office Name	Model Hospital Basaidarapur
Total Quantity	192
Item Category	Anti-SARS CoV-2/COVID-19 ELISA Testing Kit
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

### EMD Detail

Required	No
----------	----

### ePBG Detail

Required	No
----------	----

### Splitting

Bid splitting not applied.

### Anti-SARS CoV-2/COVID-19 ELISA Testing Kit ( 192 Test )

#### Technical Specifications

[\\* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL	Product Description	Anti-SARS CoV-2/COVID-19 ELISA Testing Kit	*
	Purpose	For in-vitro detection of Anti-	*

		SARS CoV-2/COVID-19 antibodies in serum/plasma of patients presenting clinical signs and symptoms or asymptomatic patients consistent with SARS CoV-2 infection	
PRODUCT INFORMATION	Type of ELISA Kit	Anti-SARS CoV-2 Human IgG ELISA Kit	*
	Detects	IgG Antibodies	*
	Detection Type	Qualitative	*
	Testing Principle	Direct ELISA	*
	Species Reactivity	Human	*
	Compatible Sample	Serum, Plasma	*
	Minimum Sensitivity	90	*
	Minimum Specificity	95	*
	Approx Result Time	Less than 3 Hours	*
	Reagents included in the kit	Conjugate, Wash buffer, Sample diluent, Sample diluent powder, TMB Substrate, Stop solution	*
	All reagents shall be in sufficient quantity to perform number of tests as per pack size	Yes	*
	Antigen Coated microwell strips provided	Yes	*
	Antigen coated strips size	96 wells (Twelve Strips with eight wells each)	*
	Positive and negative controls provided with each kit	Yes	*
	Adequate document in English detailing principle, components, methodologies, validity criteria, interpretation of result, performance characteristic, bio-safety, limitation of assay, storage condition, mfg and exp date & method of disposal provided	Yes	*
	Kit compatible with all common ELISA readers and washers	Yes	*
	Original kit literature (not photocopy) provided with each kit	Yes	*
	Storage temperature of Kit	2-8° C	*
	Shipping Condition	Under cold chain at 2-8° C	*
	The cumulative time temperature indicator technology used on each kit and prequalified by WHO	Yes	*
PACKAGING	Pack Size	96 Tests	*

ADDITIONAL REQUIREMENT	Additional Requirement	NA	*
CERTIFICATIONS & REPORTS	The kit approved from the statutory authority in its country of origin	Yes	*
	Imported kits registered and licensed in India by DCGI & Indigenous manufactured kits licensed by the authority defined under Drugs and Cosmetics act 1940 and Medical Devices Rules, 2017 as amended till date	Yes	*
	Availability of certificate/License from the Drugs Controller General of India (DCGI) for import (in case of imported kits) and marketing of the intended item	Yes	*
	Availability of valid drug license for the product issued from the competent authority defined under Drugs & Cosmetics Act, 1940 and Medical Devices Rules, 2017 as amended till date	Yes	*
	Drug license number	-	*
	Manufacturer facility certifications	ISO:13485 (Latest),GMP,WHO-GMP	*
	Product certifications/ approvals	Kit evaluated and validated by ICMR-NIV-Pune or any other ICMR validation centres	*
	Availability of validation report	Yes	*
	Availability of Test report of each batch of the product to prove the conformity to the declared specification	In house Test report from the manufacturer	*
	Submission of all required certifications, license , test reports and any other specific certification requirement as imposed by country of origin to the buyer at the time of bidding in case of bidding/ at the time of supply in case of direct order	Yes	*
SHELF LIFE	<b>Shelf life from the date of manufacture</b>	12	12, 18, 24 Or higher
	Minimum shelf life of the product at the time of delivery to the consignee	3/4 of the total shelf life	*
ADVANCE SAMPLE	Agree to provide advance sample of the product for buyer's approval before commencement of supply in	No	*

	case of bidding	
--	-----------------	--

\* Specifications highlighted in bold are the Golden Parameters.

\* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

#### **Additional Specification Documents**

#### **Consignees/Reporting Officer and Quantity**

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Vivek Kumar	110015,ESIC MODEL HOSPITAL BASAIDARAPUR NEW DELHI	192	15

#### **Bid Specific Additional Terms and Conditions**

1.Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**