



Bid Document

Bid Document			
	Bid Details		
Bid End Date/Time	27-07-2020 11:00:00		
Bid Opening Date/Time	27-07-2020 11:30:00		
Bid Life Cycle (From Publish Date)	90 (Days)		
Bid Offer Validity (From End Date)	30 (Days)		
Ministry/State Name	Jammu & Kashmir		
Department Name	Health And Family Welfare Department Jammu And Kashmir		
Organisation Name	Government Medical College		
Office Name	Govt Medical College Baramulla		
Total Quantity	2		
Item Category	ECG Machine (PAC Only)		
Bidder Turnover (Last 3 Years)	2 Lakh (s)		
OEM Average Turnover (Last 3 Years)	2 Lakh (s)		
Experience Criteria	3 Year (s)		
MSE Exemption for Years of Experience and Turnover	No		
Startup Exemption for Years of Experience and Turnover	No		
Document required from seller	Experience Criteria, Past Performance, Bidder Turnover, OEM Authorization Certificate, OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer		
Past Performance	50 %		
Bid to RA enabled	No		
Inspection Required	No		

EMD Detail

·	
Required	No

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

^{1.} Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central

/ State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

ECG Machine (2 pieces) (Under PAC)

Make	schiller
Model	cardiovitat102g2

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Certificates	Confirmity to standards	BIS	BIS
	Certification number and date	G10415050120 & Date: 17/04/2019	*
	Confirmity to safety of electromagnetic compatibility	IEC-60601-1-2:2001 (or equivalent BIS)	*
	Confirmity to safety standard	IS - 13450 / IEC60601-1-2005	*
Performance Parameters	Purpose	ECG machine is primary equipment to record ECG signal with interpretation which is required for recording and analyzing the waveforms with software.	ECG machine is primary equipment to record ECG signal with interpretation which is required for recording and analyzing the waveforms with software.
	Operating modes of ECG Machine	Automatic, Manual and Rhythm	*
	ECG machine should have ECG lead annotation facility	Yes	*
	Leads which is in ECG machine should be able to acquire simultaneously and interpret them	12	*
	Number of channels	12	12
	ECG machine should acquire lead ECG for both adult and paediatric patients	Yes	*
	The ECG machine should have facility to show lead fail indication	Yes	*

The ECG machine should have facility to show lead reversal indication	Yes	*
The ECG machine should have facility to show the impedance to quality check of connection	Yes	*
Acquisition time for ECG Machine in sec	10 sec	*
Digital sampling rate for Pacemaker spike detection	40000 s/sec/channel	*
Recording of digital sampling for pacemaker	1000 s/sec/channel	*
ECG machine should have real time colour backlit display of ECG waveforms with signal qualify indication for each lead	Yes	*
ECG machine should have frequency filters	Artifact, AC and low and high pass frequency filters	*
Number of ECGs which can be store in ECG Machine	151-200	151-200
ECG machine should have full screen preview of ECG report for quality assessment checks prior to print	Yes	Yes
Type of inbuilt screen	TFT	TFT
Size of screen in inches	7	7
Display resolution of ECG machine in pixels	640 x 480	*
ECG machine should have interpretation facility of the amplitudes, duration and morphologies of ECG waveforms and associated rhythm for adult and pediatric patient	Yes	*
ECG machine should have alphanumeric	Hard keys	Hard keys
keyboard for patient data entry		
	A4	A4
data entry	A4 200 dpi x 500 dpi on 25 mm/sec speed	A4 *
data entry Size of printing paper Resolution of digital array printer using thermal sensitive paper in ECG	200 dpi x 500 dpi on 25	
data entry Size of printing paper Resolution of digital array printer using thermal sensitive paper in ECG Machine in dpi	200 dpi x 500 dpi on 25 mm/sec speed Report formats of 3x4; 6x2, Rhythm for up to selected leads, 12 lead extended measurement, 1 minute of continuous waveform data	*

		on single charge	on single charge
	Connectivity to ECG Machine	LAN	LAN
	Storage on external portable memories	USB support	*
	The individual patient lead should be change without replacing the whole patient cable assembly	Yes	*
	Noise level in dba	45	*
	Weight of ECG Machine in Kg	4.2	*
	Frequency in Hz	300	*
	Availability of latest interpretation software	Yes	Yes
	Wireless acquisition module with RF technology	No	No
	System should have the dedicated software to download the ECG form machine in PDF format	Yes	Yes
	Warranty in years	3	3
	User and service manual of ECG Machine	Yes	*
Environmental Parameters	Operating temperature and humidity	temperature of 10-40 degree C and relative humidity of 15-90%	*
Trolley	Provision of trolley	No	No
	Material of Frame of trolley	No Trolley Provided	No Trolley Provided
	Material of Top of trolley	No Trolley Provided	No Trolley Provided
	Number of castors	No Trolley Provided	No Trolley Provided
	Number of brakes for castors	No Trolley Provided	No Trolley Provided
	Height of trolley in inch	No Trolley Provided	No Trolley Provided
	Provision of railing for top shelf	No Trolley Provided	No Trolley Provided
	Availability of suitable cable arm having holder for ECG cables when not in use	No Trolley Provided	No Trolley Provided
Power Supply	Power input	220-240 V AC, 50 Hz fitted with Indian plug	*
	Length of power cable in meter	3	*
	Resettable over current breaker shall be fitted for protection	Yes	*
Accessories	ECG Machine 12 leads with interpretation	1	1
			2

	Chest Electrodes Adult (set of Six)	1	1
	Chest Electrodes Paediatric (set of Six)	2	2
	Limb Electrodes	2 for adults and 2 for paediatric	2 for adults and 2 for paediatric
	Thermal paper for 500 patients	1 SET	1 SET
	Power cable for charging	Yes	Yes
	Supplied with Clip electrode	Yes	Yes
Reports	Manufacturer shall be ISO certified	ISO 13485	*
	ISO cetification deatils such as date and number	QS6180641505114 & Date: 26/06/2018	*
	Availability of test report from govt/ NABL/ILAC accredited lab covering all parameters	No	No
	Test report number , Incase test reports not availabble , put NA	NA	*
	Test report date, Incase test reports not availabble , put NA	NA	*
	Name of test lab , Incase test reports not availabble , put NA	NA	*
	Address of test lab , Incase test reports not availabble , put NA	NA	*
	Copies of all test reports to be provided to buyer on demand at time of supplies ,	Yes	Yes

^{*} Specifications highlighted in bold are the Golden Parameters.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1		193101,Govt Medical College Baramulla	2	15

Bid Specific Additional Terms and Conditions

- 1. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- 2.Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three

^{*} Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

This Bid is also governed by the General Terms and Conditions

---Thank You---