

Bid Document

Bid Details	
Bid End Date/Time	25-07-2020 17:00:00
Bid Opening Date/Time	25-07-2020 17:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	75 (Days)
Ministry/State Name	Ministry Of Health And Family Welfare
Department Name	Department Of Health And Family Welfare
Organisation Name	All India Institute Of Medical Sciences (aiims)
Office Name	Patna
Total Quantity	60
Item Category	Surgical Instruments - Magill Forceps
Experience Criteria	1 Year (s)
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Document required from seller	Experience Criteria,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	Yes
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

Surgical Instruments - Magill Forceps (60 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL FEATURES	Product Description	Magill Forcep	*
	Definition	A hand-held instrument used for grasping a tube [e.g., a catheter or an endotracheal (ET) tube] for its insertion and/or extraction into/from the airways, or for grasping obstructive objects for their removal from the airways. Commonly known as a Magill's forceps, it typically has a scissors-like design with ring handles and is made of high-grad stainless steel. It is available in various sizes and the working end wil typically have grasping blades that have small ringed loops or S-shaped distal working ends. The blades are serrated to provide extra grip. It is typically used by emergency medical services (EMS).	*
	Purpose	Angled forceps used to guide a tracheal tube into the larynx or a nasogastric tube into the esophagus under direct vision. It is also	*

		used to place pharyngeal packs and remove foreign bodies	
	Usage	Reusable	*
	Autoclavable	Yes	*
	Conformity to IS: 8312 (1977): Latest revision	NA	*
PRODUCT INFORMATION	Material	Surgical Stainless Steel	*
	Material Grade	Stainless Steel of 400 series (406/410/420/431)	Stainless Steel of 400 series (406/410/420/431)
	Size	Adult	Adult
	Overall Length	240	240.0
WORKMANSHIP AND FINISH	The forceps shall be free from rough edges, pits, cracks, burrs, seams and other surface	Yes	*
	All edges shall be smooth and even	Yes	*
	The serrations shall be well formed and clean	Yes	*
	The joints shall work smoothly and shall not have any undue play	Yes	*
	The forceps shall be polished bright and passivated	Yes	*
	Surface Finish	Satin Finish	*
MARKING	Each instrument shall be embossed/etched with manufacturer's name, initials or recognized trade-mark and the words 'Stainless Steel' or letters 'ss'	Yes	*
PACKAGING	Packing Type	Individual moisture proof packing	*
CERTIFICATIONS & REPORTS	Availability of Material Test Certificate from Govt authorized lab	Yes	*
	Availability of certificate from the principal company clearly mentioning location of manufacturing	Yes	*

	factories including city and country of manufacture		
	Product Certification	EU-CE (From Notified Body),US-FDA	*
	Four digit number of notified body If product is EU-CE certified	-	*
	CM/L Number if BIS Marked	-	*
	Manufacturer certification	ISO:13485 (Latest) & ISO: 9001 (Latest)	*
	Submission of all the necessary certifications eg material testing certificate, CE/US-FDA/BIS certificate, test reports and licenses to the buyer at the time of supply	Yes	*
WARRANTY	Warranty	1	1, 2, 3, 4, 5 Or higher

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Sanjeev Kumar	801507,AIIMS, Phulwarisarif, PATNA PIN-851507	60	21

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods
2. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
3. Timely Servicing / rectification of defects during warranty period: After having been notified of the defects / service requirement during warranty period, Seller has to complete the required Service / Rectification within 3 days time limit. If the Seller fails to complete service / rectification with defined time limit, a penalty of 0.5% of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG).Cumulative Penalty cannot exceed more than 10% of the total contract value after which the Buyer shall have the right to get the service / rectification done from alternate sources at the risk and cost of the Seller besides forfeiture

of PBG. Seller shall be liable to re-imburse the cost of such service / rectification to the Buyer.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---