

RA Document

RA Details	
RA Technical End Date/Time	20-07-2020 16:00:00
RA Opening Date/Time	20-07-2020 16:30:00
RA Life Cycle (From Publish Date)	90 (Days)
RA Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Delhi
Department Name	Health And Family Welfare Department Delhi
Organisation Name	N/a
Office Name	Ambedkar Nagar Hospital Dakshin Puri Delhi 110062
Total Quantity	10
Item Category	emergency recovery trolley
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

RA splitting not applied.

Emergency Recovery Trolley (10 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
---------------	--------------------	--------	----------------------------------

Trolley Dimensions	Material of Bed Frame	-	Mild Steel, Stainless Steel
	Thickness of tube of Bed frame	>= 1.2 mm	*
	Corner Bumpers	With	With, Without
	Material details of Mattress Platform in various sections	Mild Steel	*
	Details of Hydraulic Pump	-	*
	Side rail	Collapsible	Collapsible, Non Collapsible
	Height of side rail from patient Surface	360	300 - 375
	Width of X-Ray tray (mm)	-	*
	Details of Gas springs	-	*
	Overall Length of patient transfer trolley	> 2000 mm to 2100 mm	>= 1900 mm to 2000 mm, > 2000 mm to 2100 mm, >=2100 mm to 2200 mm
	Material of side rail	Stainless Steel	Stainless Steel
	Type of back rest section tilt.	Continuous Type.	Continuous Type., Step less Type
	RH and LH Fastening Belt	Without	With, Without
	Material of Fastening Belts	-	*
	Provision of Push Pull Handles; foldable handles on head side and fixed handles on Leg side of Bed Frame	Yes	*
	Overall Width of Patient Transfer Trolley (mm)	> 900 mm up to 950 mm	850 mm up to 900 mm, > 900 mm up to 950 mm Or higher
	Type of tube used in Base frame	-	*
Mattress	Type of Mattress	Single Section Mattress	Single Section Mattress
	Width of mattress (mm)	650	620 - 670
	Length of Mattress (mm) suited with trolley length	>1900 mm to 1950 mm	*
	Thickness of mattress platform	1.6 mm	1.2 mm, 1.6 mm

	Width of Mattress Platform (mm)	-	*
	Thickness of mattress	125 mm	100 mm, 125 mm
	Material of mattress	Single Layer PUF with 32 Kg/cubic Meter density	Single Layer PUF with 32 Kg/cubic Meter density, Single Layer PUF with 40 Kg/cubic Meter density, Multi Layered PUF with layer densities ≥ 32 Kg/cubic Meter
	Material of cover for mattress	-	*
	Back rest section of mattress platform shall be radiolucent X - ray	Yes	*
	Number of sections of mattress Platform (Nos)	2	2, 4
	Product Code/Drawing	Drawing Number	-
Wheels	Numbers of Castor wheel fixed on trolley	4	*
	Diameter of Castor wheel	150	125, 150 Or higher
	Numbers of centrally lockable Castor wheel (Nos)	2	1, 2 Or higher
	Numbers of swivel wheel (Nos)	2	*
	Steering facility available	Available on single side	*
	Steering facility with additional Directional Wheel	Yes	Yes, No
Trolley Height	Height Adjustable facility with dual Hydraulic pumps with foot pedals on both side of trolley	Yes	*
	Maximum Height of patient transfer trolley (mm)	≥ 840 mm to < 870 mm	*
	Range of height adjustment from the Maximum Height	300	200 - 300
Trolley Features	Adjustment of back rest section tilt	Yes	*

	Range of angle of back rest section tilt adjustment from 0 degree to maximum Degree:	80	60 - 90
	Oxygen Cylinder Holder facility	Yes	Yes
	X-ray tray assembly provided:	Yes	Yes, No
	Length of X-Ray tray (mm)	-	*
	Urine Bag Holder	Yes	Yes, No
	Range of Trendelenburg angle (+/- Degree)	15	*
	Range of Anti-Trendelenburg angle (+/- Degree)	18	0 - 20
	Facility for Knee break position	With the help of Gas Spring with Fail Safe Mechanism	Manual, With the help of Gas Spring with Fail Safe Mechanism
	Maximum angle adjustment for Knee break position (in degree)	-	*
	Load Carrying Capacity (Static) of patient transfer trolley (Kgs)	200	200.0 - 400.0 Or higher
	Load Carrying Capacity (Dynamic) of patient transfer trolley (Kgs)	200	200.0 - 400.0 Or higher
Trolley Misc Parameters	Painting of patient transfer trolley	Powder coating	*
	Colour of patient transfer trolley.	As per buyer's requirement	*
I V Pole	Telescopic I V Pole mountable on all four corners	Yes	*
	Length of I V Pole Hanging rod	-	*
	Material of I V pole	Stainless Steel	*
Warranty & Comprehensive Maintenance Contract	Warranty period	3	3, 5 Or higher
	Agreeable for AMC/CMC post warranty stage:	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by

Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignee/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days	
1	Sudhir Kumar	110062, AMBEDKAR NAGAR HOSPITAL, DAKSHIN PURI, NEW DELHI	10	2	

Bid Specific Additional Terms and Conditions

1. Provisions for workmen compensation, insurance, public liability, property damage, insurance etc. to be ensured by the contractor. Also, safety measures prescribed for working in Extra High Voltage & use of personnel protective equipment must be adhered to.

[This RA is also governed by the General Terms and Conditions](#)

---Thank You---