

Bid Document

Bid Details	
Bid End Date/Time	24-07-2020 15:00:00
Bid Opening Date/Time	24-07-2020 15:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Madhya Pradesh
Department Name	Public Health And Family Welfare Department Madhya Pradesh
Organisation Name	N/a
Office Name	451001
Total Quantity	1
Item Category	Arterial Blood Gas Analyser
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Arterial Blood Gas Analyser (1 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Certification of the product	-	*
	Certification number and date	-	*
	Conformity to Manufactures	ISO 9001 & ISO 13485	*

	Certification		
Performance Parameters	Blood Gas Analyser (BGA) System	Blood gas analyser is an equipment for measuring and calculating complete Blood Gas Analysis	*
	Type of blood gas analyser	Bench top	Bench top
	Type of technology	Cartridge (multiple use)	Cartridge (multiple use)
	Parameters measured by blood gas analyser	pCO ₂ , pO ₂ , Na ⁺ , K ⁺ , pH and calculated parameters like HCO ₃ ⁻ , Base excess, TCO ₂	*
	Additional Parameters measurable by the blood gas analyser	Cl-,iCa,Glu,BUN,Hct,Lacate, Co-oximetry,Not available	Cl-, iCa, Glu, BUN, Hct, Lacate, Co-oximetry, Not available
	Cardiac Parameters measurable	Not available	Not available
	Neonatal Parameters	nBil	nBil
	Monitor type	TFT	TFT
	Type of printer	Inbuilt Printer	Inbuilt Printer
	Cartridge / reagent to be used can be stored mostly at room temperature between 2-30 Degree C, Automatic calibration (hint: seller shall inform the exact details to buyer at the time of making the supplies)	Yes	Yes
	"Shall have patient Bar Code facility, Ability to run a wide range of patient panels on one system"	Yes	Yes
	Capability to transfer patient data to LIS/HIS	Yes	Yes
	Weight in kg	Upto 15 kg	Upto 15 kg
	Sample Volume in micro litre	65	*
	Power source	AC Mains	AC Mains
	Battery backup provided	Through UPS for 30 mins	Through UPS for 30 mins
	Shelf life of the reagent / catridges as applicable in months	3	3
	On-board shelf life of reagent in weeks	4	4
	Warranty (including electrode and membrane etc as applicable for the analyser offered) in years	3	3
	User/Technical/Maintenance manuals to be supplied in original in English in hard and soft copy	Yes	*

Reports	Demonstration of equipment and training to be provided after completing supplies before acceptance	Yes	*
	The Principal Manufacturer must have direct Presence/approved service center In India	Yes	*
	Availability of ISO certification to manufacturer	ISO 13485	*
	ISO certification date and number	-	*
	Copies of reports and certifications to be furnished to buyer on demand at time of supplies	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Narendra Joshi	451001,CMHO OFFICE, OLD HOSPITAL CAMPUS, KHARGONE	1	15

Bid Specific Additional Terms and Conditions

1. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---