

Bid Document

| Bid Details | |
|--|----------------------------|
| Bid End Date/Time | 22-07-2020 19:00:00 |
| Bid Opening Date/Time | 22-07-2020 19:30:00 |
| Bid Life Cycle (From Publish Date) | 90 (Days) |
| Bid Offer Validity (From End Date) | 15 (Days) |
| Ministry/State Name | Ministry Of Railways |
| Department Name | Na |
| Organisation Name | N/a |
| Office Name | Eastern Railway |
| Total Quantity | 800 |
| Item Category | Surgical Drapes-Disposable |
| MSE Exemption for Years of Experience and Turnover | No |
| Startup Exemption for Years of Experience and Turnover | No |
| Bid to RA enabled | No |
| Inspection Required | No |

EMD Detail

| | |
|----------|----|
| Required | No |
|----------|----|

ePBG Detail

| | |
|----------|----|
| Required | No |
|----------|----|

Splitting

Bid splitting not applied.

Surgical Drapes-Disposable (800 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|---------------|------------------------|------------------------|----------------------------------|
| GENERAL | Type of Surgical Drape | General Surgery Drapes | General Surgery Drapes |
| | Name of Procedure | General Surgery Drape | General Surgery Drape |

| | | | |
|------------|-------------------------------------|--|---------------------------------------|
| | Color | Medical Blue | * |
| | Viral & Bacterial Barrier | Yes | * |
| | Tear and Puncture Resistant | Yes | * |
| | Loose Fibre and Lint Free | Yes | * |
| | Impervious to Fluid | Yes | * |
| | Chemical Free | Yes | * |
| | Non Drug | Yes | * |
| | Disposable | Yes | * |
| | Breathable | Yes | * |
| | Incise Film | No | No |
| | Leggings Attached | No | No |
| | Lithotomy Position | No | No |
| | Finger Guard | No | No |
| | Drainage Pouch | No | No |
| | Drainage Tube | No | No |
| | Sterilized | Yes | Yes |
| | Method of Sterilization | Other | * |
| | Adhesive Strip | No | No |
| DIMENSIONS | Length of drape | 120 | 120.0 |
| | Width of Drape | 100 | 100.0 |
| | Length of incise film | 7 | 7.0 |
| | Width of incise film | 60 | 60.0 |
| | Drain Pouch Length | 35 | 35.0 |
| | Drain Pouch Width | 75 | 75.0 |
| | Drainage Tube Length (cm) | 25 | 25.0 |
| | Drainage Tube Width (cm) | 2 | 2.0 |
| | Adhesive Strip width (cm) | 10 | 10.0 |
| | Finger guard length (cm) | 20 | 20.0 |
| | Finger guard dia (cm) | 10 | 10.0 |
| MATERIAL | Fabric Details of Drape | Non Woven, Reinforced | Non Woven, Reinforced |
| | Fabric Composition Details | Spunbond + meltblown + Spunbond (SMS) | Spunbond + meltblown + Spunbond (SMS) |
| | GSM of the Fabric (g/m2) | 42 | 42 |
| | Material of incise film | NA | * |
| | Material of Drain Pouch/Tube | Non-Woven, Polycoated | Non-Woven, Polycoated |
| | Fabric for Fenestration of Drape | Reinforcement of Polycoated fabric, adsorbent on one side and impervious to fluids on other around fenestration area | * |
| | Material of Finger Guard | NA | * |

| | | | |
|----------------|---|---|---------|
| CONSTRUCTION | Drape Coated with Pressure Sensitive Acrylate Adhesive | NA | * |
| | Medical grade adhesive layer protected with peel off thick silicon coated release paper | NA | * |
| | Double side adhesive tapes on all corners and central edges of sheets | NA | * |
| PACKING | Type of Packing | Sterile Pack, must be double packed with outer pouch must be peel off made of Medical Grade Paper | * |
| CERTIFICATIONS | AAMI Protection Level | Level I | Level I |
| | Certification | - | * |
| | Submission of Test Report clearly specifying the nomenclature of fabric, method of Sterilization and AAMI Protection level in addition to other parameters if any, to the buyer on Form39 or from NABL/ILAC accredited or Central Government Lab to prove conformity of products to the specification | Yes | * |
| | Product Conformity certificate to be furnished to buyer at the time of Supply | Yes | * |

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporting Officer | Address | Quantity | Delivery Days |
|-------|-----------------------------|---|----------|---------------|
| 1 | Bidyut Kumar Mondal | 713301,DRM Building, Eastern Railway, Asansol | 800 | 15 |

Bid Specific Additional Terms and Conditions

- 1.Scope of supply (Bid price to include all cost components) : Only supply of Goods
- 2.IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
- 3.Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
- 4.Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
- 5.Warranty period of the supplied products shall be 1 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---