

RA Number: GEM/2020/RA/43102

Dated: 17-06-2020

#### **RA Document**

KA Document				
RA Details				
RA Technical End Date/Time	27-06-2020 18:00:00			
RA Opening Date/Time	27-06-2020 18:30:00			
RA Life Cycle (From Publish Date)	90 (Days)			
RA Offer Validity (From End Date)	30 (Days)			
Ministry/State Name	Telangana			
Department Name	Animal Husbandry And Fisheries Department Telangana			
Organisation Name	N/a			
Office Name	Shanthinagar,hyderabad			
Total Quantity	1			
Item Category	Laboratory Incubators			
MSE Exemption for Years of Experience and Turnover	No			
Startup Exemption for Years of Experience and Turnover	No			
Document required from seller	OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer			

#### **EMD Detail**

Required	No

#### ePBG Detail

Required	No
•	

## Splitting

RA splitting not applied.

# Laboratory Incubators ( 1 pieces )

#### **Technical Specifications**

### \* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	-	*
	Certification number and date	-	*
	Confirmity to standard for for electrical safety latest amendment	IEC - 61010	*
Performance Parameters	Purpose	Bacteriological incubator are used for the incubation of biological products under controlled conditions	*
	Type of Incubator	Bacteriological Incubator	Bacteriological Incubator
	Capacity in liters	201-300	201-300
	Material of Inner Chambers	SS 304	*
	Material of Outer Chamber	CRCA (powder coated point) Steel	CRCA (powder coated point) Steel
	Ambient Temperature in °C	5°C above ambient to 60 °C (Bacteriological Incubator)	5°C above ambient to 60 °C (Bacteriological Incubator)
	Temperature Accuracy in °C	±0.5°C	*
	Temperature Uniformity in °C	±2°C	*
	Insulation	Glass wool	Glass wool
	Controller	Microprocessor based PID control	*
	Digital Display of temperature in °C	Yes	*
	Type of Display	LED	LED
	Type of Temperature sensor	PT100	*
	Power supply	220 V , 50 Hz Single Phase	*
	Door	Insulated door and fitted with heavy hinges and should have double glass window	*
	Type of Shelves	Removable	Removable

	Material of Shelves	SS wire mesh	*
	Number of shelves	3	3
	Size of incubator (L x W x H) in mm x mm x mm	-	*
	Weight in kgs	101	*
	Auto power break off	Yes	Yes
	Over temperature protection	Yes	*
	Safety alarm system	Yes	*
	Instruction manual to be provided	Yes	*
	Warranty in Years	1	1
•	Manufacturer should have ISO certification	Yes	*
	ISO certification date and number	-	*
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*

<sup>\*</sup> Specifications highlighted in bold are the Golden Parameters.

#### **Additional Specification Documents**

#### **Consignee/Reporting Officer and Quantity**

S.No.	Consignee/Reporti ng Officer	Address	Quantity	Delivery Days	
1	M Rama Surya Narasimha Neehar	500028,D.No.10-2-289-127, Shanthinagar	1	15	N/A

## **Bid Specific Additional Terms and Conditions**

- 1. Scope of supply (Bid price to include all cost components): Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
- 2. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- 3. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

<sup>\*</sup> Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

4. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.

This RA is also governed by the General Terms and Conditions

---Thank You---