

Dated: 10-07-2020



Bid Document

| | Document | | | |
|--|---|--|--|--|
| Bid Details | | | | |
| Bid End Date/Time | 22-07-2020 17:00:00 | | | |
| Bid Opening Date/Time | 22-07-2020 17:30:00 | | | |
| Bid Life Cycle (From Publish Date) | 90 (Days) | | | |
| Bid Offer Validity (From End Date) | 45 (Days) | | | |
| Ministry/State Name | Ministry Of Labour And Employment | | | |
| Department Name | Na | | | |
| Organisation Name | Employees State Insurance Corporation (esic) | | | |
| Office Name | Esic Medical College & Hospital, Nh-3, Faridabad | | | |
| Total Quantity | 1 | | | |
| Item Category | Deep freezer | | | |
| Bidder Turnover (Last 3 Years) | 7 Lakh (s) | | | |
| OEM Average Turnover (Last 3 Years) | 56 Lakh (s) | | | |
| Experience Criteria | 3 Year (s) | | | |
| MSE Exemption for Years Of Experience and Turnover | Yes | | | |
| Startup Exemption for Years Of Experience and Turnover | Yes | | | |
| Document required from seller | Experience Criteria, Past Performance, Bidder Turnover, Certificate (Requested in ATC), OEM Authorization Certificate, OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer | | | |
| Past Performance | 50 % | | | |
| Bid to RA enabled | Yes | | | |
| Inspection Required | No | | | |
| Estimated Bid Value | 1400000 | | | |
| | | | | |

EMD Detail

| Advisory Bank | State Bank of India | |
|-------------------|---------------------|--|
| EMD Percentage(%) | 1.00 | |
| EMD Amount | 14000 | |

ePBG Detail

| Advisory Bank | State Bank of India | |
|-------------------------------------|---------------------|--|
| ePBG Percentage(%) | 5.00 | |
| Duration of ePBG required (Months). | 62 | |

- (a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.
- (b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

Beneficiary:

Dean

Esic Medical College & Hospital, Nh-3, Faridabad, NA, Employees State Insurance Corporation (ESIC), Ministry of Labour and Employment (Dr. Asim Das)

Splitting

Bid splitting not applied.

- 1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
- 2. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
- 3. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 50% of bid quantity, in at least one of the last three years before the bid opening date to any Central / State Govt Organization / PSU / Public Listed Company. Copies of relevant contracts (proving supply of cumulative order quantity in any one year) to be submitted along with bid in support of quantity supplied in the relevant year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

Deep Freezer (1 pieces)

Technical Specifications

* As per GeM Category Specification

| Specification | Specification Name | Values | Bid Requirement (Allowed Values) |
|---------------------------|--|---|--|
| Standards | Conformity to Standards | CE (with 4 digit notified number) | CE (with 4 digit notified number), USFDA |
| | Certification date and number | - | * |
| | Conformity to electrical safety | IEC 61010-1, IEC 61010-2-40 or IS 13450 | * |
| Performance Parameters | Purpose | For storage of various biological products including ATCC cultures, enzymes, vaccines, chemicals or material testing components for a longer period of time | * |
| | Capacity in liters | 501-600 | 501-600 |
| | Type of cabinet | Vertical | Vertical |
| | Ground Clearance in mm | 100 mm | * |
| | Cooling type | Direct cooling | * |
| | Freezer body | Galvanised steel body with tough powder coated exterior finish | * |
| | Castors | Heavy duty lockable | Heavy duty lockable |
| | Lockable outer doors and lids | Yes | * |
| | Material of internal cabinet | SS | * |
| | Freezer should use CFC- free, HCFC-free non flammable refrigerants | Yes | * |
| | Refrigeration system must be energy efficient | Yes | * |
| | Pull down time in hrs | 5.2 hrs | * |
| | Operating Temperature in degree celcius | minus 10 degree to minus 40 | * |
| | Increment variation setting | 1° C | * |
| | Number of corrosion resistant racks | 4 | 4 |

| | Control | Fully programmable microprosseor controlled with membrane keypad and eye level control panel | * |
|-----------------------|--|--|-----|
| | Number of inner stroage compartments with insulated doors and | 4 | 4 |
| | Digital temperature display | Yes | * |
| | Should have battery back up for the display and security lock for the display | Yes | * |
| | Alarms | Audible and visible alarm system for unwanted temperature rise, sudden power failure, system failure and battery low | * |
| | Noise level in dB | 60 | * |
| | Energy consumption rate (@KWh/day) | 8.3 | * |
| | Stabilizer should be capable to run any voltage between 190V - 270 V | Yes | * |
| | Warranty of Stabilizer in Years | 5 years | * |
| | Warranty of Freezer in Years (from the date of installation) | 5 | 5 |
| Additional Parameters | Freezer must have washable condenser filter indicator which should keep fins free of dust to maintain peak cooling efficiency | Yes | * |
| | One 5KVA servo stabilizer to be supplied with freezer | Yes | * |
| | CO ₂ cylinder should be supplied with freezer for backup | Yes | Yes |
| | Tempearture Chart recorder to be provided | Yes | Yes |
| Reports | Manufacturer should have ISO certification | Yes | * |

| ISO certification date and number | - | * |
|--|-----|---|
| Copies of all certifications and reports to be provided to buyer on demand at time of supplies | Yes | * |

^{*} Specifications highlighted in bold are the Golden Parameters.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

| S.No. | Consignee/Reporti ng Officer | Address | Quantity | Delivery Days |
|-------|---------------------------------|--|----------|---------------|
| 1 | Harsh Nandini Arora | 121001,ESIC Medical College and hospital NH-3, NIT Faridabad | 1 | 15 |

Bid Specific Additional Terms and Conditions

- 1. IMPORTED PRODUCTS: In case of imported products, OEM or Authorized Seller of OEM should have a registered office in India to provide after sales service support in India. The certificate to this effect should be submitted.
- 2. Successful Bidder can submit the Performance Security in the form of Fixed Deposit Receipt also (besides PBG which is allowed as per GeM GTC). FDR should be made out or pledged in the name of The Dean, ESIC Medical College & Hospital, NH-3, NIT Faridabad A/C (Name of the Seller). The bank should certify on it that the deposit can be withdrawn only on the demand or with the sanction of the pledgee. For release of Security Deposit, the FDR will be released in favour of bidder by the Buyer after making endorsement on the back of the FDR duly signed and stamped along with covering letter. Successful Bidder has to upload scanned copy of the FDR document in place of PBG and has to ensure delivery of hard copy of Original FDR to the Buyer within 15 days of award of contract.
- 3. Scope of supply (Bid price to include all cost components): Supply Installation Testing Commissioning of Goods and Training of operators and providing Statutory Clearances required (if any)
- 4. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
- 5. ISO 9001: The bidder must have ISO 9001 certification.
- 6. The bidder is required to upload, along with the bid, all relevant certificates such as BIS licence, type test certificate, approval certificates and other certificates as prescribed in the Product Specification given in the bid document.
- 7. Bidder Turn Over Criteria: The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.

^{*} Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

- 8. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria. In case of bunch bids, the OEM of CATEGORY RELATED TO primary product having highest bid value should meet this criterion.
- 9. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- 10. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.
- 11. Escalation Matrix For Service Support : Bidder/OEM must provide Escalation Matrix of Telephone Numbers for Service Support.
- 12. Warranty period of the supplied products shall be 5 years from the date of final acceptance of goods or after completion of installation, commissioning & testing of goods (if included in the scope of supply), at consignee location. OEM Warranty certificates must be submitted by Successful Bidder at the time of delivery of Goods. The seller should guarantee the rectification of goods in case of any break down during the guarantee period. Seller should have well established Installation, Commissioning, Training, Troubleshooting and Maintenance Service group in INDIA for attending the after sales service. Details of Service Centres near consignee destinations are to be uploaded along with the bid
- 13. The Bidder / OEM {themselves or through reseller(s)}, should have executed project for supply and installation / commissioning of same or similar Category Products during preceding 3 financial years (i.e. current year and three previous financial years) as on opening of bid, as per following criteria:
 - (i) Single order of at least 35% of estimated bid value; or
 - (ii) Two orders of at least 20% each of estimated bid value; or
 - (iii) Three orders of at least 15% each of estimated bid value.
 - Satisfactory Performance certificate issued by respective Buyer Organization for the above Orders should be uploaded with bid. In case of bunch bids, the Category related to primary product having highest bid value should meet this criterion
- 14. For fulfilling the experience criteria any one of the following documents may be considered as valid proof for meeting the experience criteria:
 - a. Purchase Order copy along with Invoice(s) with self-certification by the bidder that supplies against the invoices have been executed.
 - b. Execution certificate by client with order value.
 - c. Any other document in support of order execution like Third Party Inspection release note, etc.

This Bid is also governed by the General Terms and Conditions

---Thank You---