

Bid Document

Bid Details	
Bid End Date/Time	23-07-2020 13:00:00
Bid Opening Date/Time	23-07-2020 13:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	80 (Days)
Ministry/State Name	Ministry Of Labour And Employment
Department Name	Employees State Insurance Corporation
Organisation Name	Employees State Insurance Corporation
Office Name	Model Hospital Basaidarapur
Total Quantity	80
Item Category	Medical X Ray Films
MSE Exemption for Years of Experience and Turnover	No
Startup Exemption for Years of Experience and Turnover	No
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

Medical X Ray Films (80 packet)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	IS 15584:2005	*
	BIS Marked	No	Yes, No
	CM/L Number to be provided if not BIS marked put NA	-	*
Performance Parameters	Purpose	Double sided emulsion coated medical x ray films used for general purpose radiography IN SHEET FORM	*
	Sensitivity to type of light	Blue sensitive	Blue sensitive
	Material of the base	Polyethylene terephthalate (PET) resin with blue Tone	Polyethylene terephthalate (PET) resin with blue Tone, Polyethylene terephthalate (PET) resin with clear tone
	Base density	0.10±0.01	*
	Films shall be free from visual and latent defects	Yes	*
	Conforming specification for size of x ray films	IS 9644	*
	Sizes of the films in mm x mm	356 x 356	356 x 356
	Thickness of the base in microns with tolerance ±20	185	*
	Maximum density D Max,Min	2.25	*
	Speed which is reciprocal of air kerma in gray required to produce a density of 1 above base plus fog	1.6	*
	Average gradient (G)	2.0-3.0	*
	The four corners of the cut sheets shall be rounded and they shall not have any steps and sharp features	Yes	*

Packing Parameters	Interleaving paper for the protection of the film to be provided	Yes	*
	Minimum Weight of the interleaved paper	50 GSM	*
	Colour of the interleaving paper	White	White, Yellow
	Number of sheets of films for medical X Ray	50	50
	Sheets shall be placed in light proof black pouches made of low density polyethylene film of at least 175µ thick	Yes	*
	The pouches of sheets shall be put in light tight cartons and surely sealed with adhesive tape	Yes	*
	Guaranteed shelf life from date of manufacture	24 months	18 months, 24 months
	At least 2/3 of shelf life to be available at time of supply	Yes	*
	Marking on the carton must include Trade mark / name of the manufacturer, size of the film, number of sheets, date of expiry, blue/green sensitive film AND SPEED classification and screen to be used	Yes	*
Reports	ISO certifications for Manufacturer	Yes	*
	ISO certification number and date	-	*
	Availability of test report from central govt/ NABL/ILAC accredited lab covering all parameters	No	Yes, No
	Test report number put NA if test report not available	-	*

Test report date put NA if test report not available	-	*
Name of test lab put NA if test report not available	-	*
Address of test lab put NA if test report not available	-	*
Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Vivek Kumar	110015,ESIC MODEL HOSPITAL BASAIDARAPUR NEW DELHI	80	15

Bid Specific Additional Terms and Conditions

1. Scope of supply (Bid price to include all cost components) : Only supply of Goods

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---