

Bid Document

Bid Details	
Bid End Date/Time	25-07-2020 12:00:00
Bid Opening Date/Time	25-07-2020 12:30:00
Bid Life Cycle (From Publish Date)	90 (Days)
Bid Offer Validity (From End Date)	30 (Days)
Ministry/State Name	Gujarat
Department Name	Health & Family Welfare Department Gujarat
Organisation Name	N/a
Office Name	Indian System Of Medicine And Homeopathy
Total Quantity	3
Item Category	Steam Sterlizer
Experience Criteria	1 Year (s)
MSE Exemption for Years Of Experience	Yes
Startup Exemption for Years Of Experience	Yes
Document required from seller	Experience Criteria,OEM Authorization Certificate *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer
Bid to RA enabled	No
Inspection Required	No

EMD Detail

Required	No
----------	----

ePBG Detail

Required	No
----------	----

Splitting

Bid splitting not applied.

1. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM {themselves or through reseller(s)} should have regularly, manufactured and supplied same or

similar Category Products to any Central / State Govt Organization / PSU / Public Listed Company for number of years as indicated in the bid document before the bid opening date. Copies of relevant contracts to be submitted along with bid in support of having supplied some quantity during each of the year. In case of bunch bids, the category of primary product having highest value should meet this criterion.

Steam Sterlizer (3 pieces)

Technical Specifications

[* As per GeM Category Specification](#)

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Construction	Type of Steriliser	HORIZONTAL CYLINDRICAL Steam Steriliser Pressure type for Hospital and Pharmaceutical use, Steriliser Chamber size: 500 mm (Diameter) x 900 mm (Length), conforming to IS: 3829(part 1)/1999, fitted with Silicon door gasket (expansion type). Electric load load of the unit shall be 9 KW	HORIZONTAL CYLINDRICAL Steam Steriliser Pressure type for Hospital and Pharmaceutical use, Steriliser Chamber size: 500 mm (Diameter) x 900 mm (Length), conforming to IS: 3829(part 1)/1999, fitted with Silicon door gasket (expansion type). Electric load load of the unit shall be 9 KW
	Steam supply to Steriliser	Direct Steam Operation from an external source	Direct Steam Operation from an external source
	Material: Steriliser parts; Chamber, Jacket, Door, Door Ring and shall be made of	SS 316 or SS 316L	SS 316 or SS 316L
	Type of control	Automatic: MICROPROCESSOR CONTROLLED, WITH PRESELECTED STERILISATION PROGRAMS (variable programs customisable as per Buyer's requirement to deal with sterilisation and drying of wrapped/unwrapped loads at 121 & 134 degree C)	Automatic: MICROPROCESSOR CONTROLLED, WITH PRESELECTED STERILISATION PROGRAMS (variable programs customisable as per Buyer's requirement to deal with sterilisation and drying of wrapped/unwrapped loads at 121 & 134 degree C)

CONSTRUCTION	Instruction plate on Steam Steriliser advising user to use soft water for steam generation for long service life of Steam Steriliser	YES	*
	Whether a compatible water softener/reversed osmosis based system to feed water to the steam generator unit is to be supplied with the steriliser:	No	No
	Provision of Lock on Steriliser Chamber door:	Manual Lock	Manual Lock
	Installation	The site ready for installation equipped with water supply, electric connection cabling and drainage outlet shall be provided by Buyer. The Seller shall be responsible to carry out for installation of the offered Steam Steriliser unit on turnkey basis i.e. Any civil work required for installation of the ordered equipment i.e. flooring/ tilling/plastering/Paint and/or wood work shall be Seller's responsibility. The Seller shall also provide training to personnel of buyer.	*
	Type of Door:	Sliding type	Sliding type
	Display facility for Sterilisation Cycle parameters	Chamber Temperature	Chamber Temperature
	Print facility for record generation of dates, time, load, identification number and operating parameters (temperature and	Yes, port given to connect an external Printer Unit	Yes, port given to connect an external Printer Unit

	holding time etc):		
	Warranty Period	3	3, 5 Or higher
	No of Doors in Steriliser	1	1
	Whether Type Test Certificate from a Govt/NABL/ILAC approved testing laboratory is available (If Yes, to be furnished by Seller to Buyer on demand) for the offered product?	No	No
	Type of user friendly Interface (for microprocessor controlled Sterilizer) to specify parameters for running Sterilization cycle:	NA	NA
	Data Transfer facility through USB /RS232 port to computer/laptop on real time basis	No	No
	Air removal from Sterilisation chamber, after loading the steriliser, by on for more pressure/vacuums pulsing for proper steam penetration	YES	*
Integrated Shredder Facility	Integrated Shredder (for shredding the load in Sterilisation chamber)	Not available	Not available
	Material: Shredder parts;	SS 304	SS 304, SS 316 or SS 316L
	Power of Motor used for Shredding	NA, no integrated Shredder unit offered	NA, no integrated Shredder unit offered
Accessories (made of SS 316) for loading/unloading and handling of Load	Number of Rails:	0	0, 2 Or higher
	Long Handle:	0	0
	Sterilisation Carriages	0	0, 1, 2 Or higher
	Pull out Trays/Tanks	0	0
	Floor loading carts	0	0, 2 Or higher

Transfer carriages:	0	0, 1, 2 Or higher
----------------------------	---	-------------------

* Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

Additional Specification Documents

Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Kamlesh Kanjibhai Chaudhary	384151,GOVT. HOMIOPETHY COLLEGE & HOSPITAL, DETHALI, TA-SIDDHPUR, DIST-PATAN PINCODE-384151	3	15

Bid Specific Additional Terms and Conditions

1. Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
2. **Upload Manufacturer authorization:** Wherever Authorised Distributors are submitting the bid, Manufacturers Authorisation Form (MAF)/Certificate with OEM details such as name, designation, address, e-mail Id and Phone No. required to be furnished along with the bid.
3. Bidder's offer is liable to be rejected if they don't upload any of the certificates / documents sought in the Bid document, ATC and Corrigendum if any.
4. Dedicated /toll Free Telephone No. for Service Support : BIDDER/OEM must have Dedicated/toll Free Telephone No. for Service Support.

[This Bid is also governed by the General Terms and Conditions](#)

---Thank You---