



Bid Document

Bid Document				
Bid Details				
24-07-2020 11:00:00				
24-07-2020 11:30:00				
90 (Days)				
30 (Days)				
Ministry Of Defence				
Department Of Defence				
Indian Army				

2000				
surgical gloves				
No				

EMD Detail

Ι.		
Ш		
	Required	No
	Required	140

ePBG Detail

Required No	
---------------	--

Splitting

Bid splitting not applied.

Surgical Gloves (2000 pairs)

Technical Specifications

* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
GENERAL	Type of Surgical Glove	Surgical Rubber Gloves conforming to IS 4148 (1989)	Disposable surgical or Gynaecological rubber gloves conforming to IS

			13422 (1992)-powder free, Surgical Rubber Gloves conforming to IS 4148 (1989)
	Sterlity	Non- sterlite	Sterlized, Non- sterlite
	Finish of outer surface	Smooth	*
	Colour	Translucent	*
	Colouring Agent	No	*
	Use	Multiple	Single, Multiple
	Shelf Life in years	4	*
	The product should have at least 2/3 rd of the total shelf life available at the time of supply	Yes	*
STANDARDS	Conformity to Indian Pharmacopieia for Sterility (if Sterlized)	NA	*
DIMENSIONS	Size (Number)	8.0 ½	5.0 ½, 6, 6.0 ½, 7, 7 .0 ½, 8, 8.0 ½, 9, 9.0 ½
	Thickness	0.24	0.1, 0.24
MATERIAL	Material	Synthetic Rubber Latex	Natural Rubber Latex, Synthetic Rubber Latex, Rubber Cement
	Lubricant / Material used for surface treatment	-	*
	Powdered	No	*
	PH of aqueous extract from rubber	6	*
PACKING	Packed to maintain sterlity during shipping & storage and permit opening without contamination of the gloves	NA	YES, NA
	No. of Gloves in a pack	2	*
MARKING	ISI Marked	YES	*
	CM/ L No.	-	*
REPORTS	Submission of Test Reports on Form 39 or from Central Govt./NABL/IL AC acredited Lab to prove conformity to the specifications at the time of supply	-	*

Additional Specification Documents

Consignees/Reporting Officer and Quantity

	S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
- 1					

^{*} Specifications highlighted in bold are the Golden Parameters.

* Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement (allowed Values) by the Buyer.

1	******	********DEHRADUN	2000	7	
Bid Specific Additional Terms and Conditions					
1.Scope of supply (Bid price to include all cost components) : Only supply of Goods					
This Bid is also governed by the General Terms and Conditions					
Thank You					