



#### **Bid Document**

Dia Document			
Bid Details			
Bid End Date/Time	24-07-2020 11:00:00		
Bid Opening Date/Time	24-07-2020 11:30:00		
Bid Life Cycle (From Publish Date)	90 (Days)		
Bid Offer Validity (From End Date)	80 (Days)		
Ministry/State Name	Madhya Pradesh		
Department Name	Home Department Madhya Pradesh		
Organisation Name	N/a		
Office Name	State Forensic Science Lab Sagar		
Total Quantity	19		
Item Category	Laboratory Refrigerator		
MSE Exemption for Years of Experience and Turnover	No		
Startup Exemption for Years of Experience and Turnover	No		
Bid to RA enabled	No		
Inspection Required	No		

## **EMD Detail**

Advisory Bank	State Bank of India
EMD Percentage(%)	0.50
EMD Amount	3297

#### ePBG Detail

	<u> </u>
I I Doguirod	l No
i i Reduired	I No

(a). EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

## **Beneficiary:**

Store Officer

STATE FORENSIC SCIENCE LAB SAGAR, Home Department Madhya Pradesh, N/A, (Sushil Kumar Raikwar)

## **Splitting**

Bid splitting not applied.

# Laboratory Refrigerator ( 19 pieces )

## **Technical Specifications**

## \* As per GeM Category Specification

Specification	Specification Name	Values	Bid Requirement (Allowed Values)
Standards	Conformity to Standards	CE (with 4 digit notified number)	CE (with 4 digit notified number)
	Certification date and number	-	*
	Electrical safety standards	IEC 61010 With latest amendments	IEC 61010 With latest amendments
Performance Parameters	Purpose	The laboratory Refrigerators are specifically designed and are suitable for research laboratories, Medicine storage, blood banks applications and across other laboratories working with critical elements	*
	Capacity of the refrigerator in liters	340-360	340-360
	Temperature range in °C	1°C to 8°C throughout the chamber	*
	Control panel settings	Thermometer ,Main switch and temp selection	*
	Temperature Control	Microprocessor based temperature controller	*
	Type of Door	Single door	Single door
	Material of Construction of body	Galvanized steel	*
	Insulation material	High grade foam material	*
	Outside finish	Epoxy coated finish	*
	Should have digital display of temperature	Yes	*
	Rotary air circulation to maintain temperature uniformity	Yes	*
	Regrigerator should be frost free and CFC free	Yes	*
	Flouroscent internal light available	Yes	*
	Power supply	220-240 V , 50 Hz Single phase	*
	Door locking system for improved security with magnetic door gasket	Yes	*
	Refrigerator Energy star rating	3	3

	Material used for shelves	Toughened glass	*
	Number of shelves	4	4
	Refrigerator shall be hermentically sealed	Yes	*
	Voltage stabilizer to be supplied	Yes	*
	Material of Inner chamber	SS 304	*
	Warranty of the refrigerator in years	3	3
	Controller based audiovisual alarm	No	No
	Compressor shall be low noise and vibration proof	Yes	*
	Warranty of Compressor in years	5	5
Reports	Manufacturer should have ISO certification	Yes	*
	ISO certification date and number	-	*
	Availability of test report from central govt/ NABL/ILAC accredited lab covering all parameters	Yes	*
	Test report date	-	*
	Test report number	-	*
	Name of the Lab	-	*
	Address of the Lab	-	*
	Copies of all certifications and reports to be provided to buyer on demand at time of supplies	Yes	*

<sup>\*</sup> Specifications highlighted in bold are the Golden Parameters.

#### **Additional Specification Documents**

## **Consignees/Reporting Officer and Quantity**

	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Jyotsna Pandey	470001,STATE FORENSIC SCIENCE LAB. POLICE LINES SAGAR (M.P)	19	15

## **Bid Specific Additional Terms and Conditions**

- 1.Availability of Service Centres: Bidder/OEM must have a Functional Service Centre in the State of each Consignee's Location in case of carry-in warranty. (Not applicable in case of goods having on-site warranty). If service center is not already there at the time of bidding, successful bidder / OEM shall have to establish one within 30 days of award of contract. Payment shall be released only after submission of documentary evidence of having Functional Service Centre.
- 2.Procurement under this bid is reserved for purchase from Micro and Small Enterprises from the State of Bid Inviting Authority whose credentials are validated online through Udyog Aadhaar for that product category. If the bidder wants to avail the reservation benefit, the bidder must be the manufacturer of the offered product in case

<sup>\*</sup> Bidders may note that In respect of non-golden Parameters, the specifications 'Values' chosen by Buyer will generally be preferred over 'Bid requirement ( allowed Values) by the Buyer.

- of bid for supply of goods. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service.
- 3.Successful bidder will have to ensure that adequate number of dedicated technical service personals / engineers are designated / deployed for attending to the Service Request in a time bound manner and for ensuring Timely Servicing / rectification of defects during warranty period, as per Service level agreement indicated in the relevant clause of the bid.

This Bid is also governed by the General Terms and Conditions

---Thank You---